



IEC 60601-2-16

Edition 5.0 2018-04
REDLINE VERSION

INTERNATIONAL STANDARD



**Medical electrical equipment –
Part 2-16: Particular requirements for the basic safety and essential
performance of haemodialysis, haemodiafiltration and haemofiltration
equipment**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.20; 11.040.25

ISBN 978-2-8322-5624-4

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	2
201.1 Scope, object and related standards.....	8
201.2 Normative references	10
201.3 Terms and definitions	11
201.4 General requirements	14
201.5 General requirements for testing ME EQUIPMENT	18
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	18
201.7 ME EQUIPMENT identification, marking and documents	18
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	22
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	23
201.10 Protection against unwanted and excessive radiation HAZARDS	23
201.11 Protection against excessive temperatures and other HAZARDS	24
201.12 * Accuracy of controls and instruments and protection against hazardous outputs	25
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	35
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	36
201.15 Construction of ME EQUIPMENT.....	36
201.16 * ME SYSTEMS	37
201.17 ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS	38
202 Electromagnetic compatibility disturbances – Requirements and tests	38
208 General requirements, tests and guidance for ALARM SYSTEMS in MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS	39
209 Requirements for environmentally conscious design	40
210 Process Requirements for the development of PHYSIOLOGIC CLOSED-LOOP CONTROLLERS	41
211 * Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the HOME HEALTHCARE ENVIRONMENT.....	41
Annexes	43
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures.....	44
Annex AA (informative) Particular guidance and rationale.....	45
Annex BB (informative) Examples of HAZARDS, foreseeable sequences of events, and HAZARDOUS SITUATIONS in HAEMODIALYSIS EQUIPMENT	66
Bibliography.....	75
Index of defined terms used in this particular standard.....	78
Figure 201.101 – Continuous air infusion test setup with example dimensions	32
Figure AA.1 – Example of a HAEMODIALYSIS ME SYSTEM	62
Table 201.101 – ESSENTIAL PERFORMANCE requirements	14

Table AA.1 – ~~Possible~~ Example of ALARM CONDITION priorities according to 6.1.2 of IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, adapted for HAEMODIALYSIS EQUIPMENT needs..... 64

Table BB.1 – HAZARDOUS SITUATION list following ISO 14971:2007, Annex E 66

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-16: Particular requirements for the basic safety and essential performance of haemodialysis, haemodiafiltration and haemofiltration equipment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This redline version of the official IEC Standard allows the user to identify the changes made to the previous edition. A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text.

International standard IEC 60601-2-16 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This fifth edition cancels and replaces the fourth edition of IEC 60601-2-16 published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) update of references to IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, of references and requirements to IEC 60601-1-2:2014, of references to IEC 60601-1-6:2010 and IEC 60601-1-6:2010/AMD1:2013, of references and requirements to IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, of references to IEC 60601-1-9:2007 and IEC 60601-1-9:2007/AMD1:2013, of references to IEC 60601-1-10:2007 and IEC 60601-1-10:2007/AMD1:2013 and of references to IEC 60601-1-11:2015;
- b) widening of the scope;
- c) editorial improvements;
- d) addition of requirements for anticoagulant delivery means;
- e) other few small technical changes.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/1557/FDIS	62D/1585/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION EQUIPMENT.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-16: Particular requirements for the basic safety and essential performance of haemodialysis, haemodiafiltration and haemofiltration equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Addition Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION EQUIPMENT, hereafter referred to as HAEMODIALYSIS EQUIPMENT.

This document does not take into consideration **specific safety details** of the DIALYSIS FLUID control system of HAEMODIALYSIS EQUIPMENT using regeneration of DIALYSIS FLUID ~~and or~~ CENTRAL DELIVERY SYSTEMS **for DIALYSIS FLUID**. It does, however, take into consideration the specific safety requirements of such HAEMODIALYSIS EQUIPMENT concerning electrical safety and PATIENT safety.

This document specifies the minimum safety requirements for HAEMODIALYSIS EQUIPMENT. These ~~devices~~ **HAEMODIALYSIS EQUIPMENT** are intended for use either by medical staff or for use by the PATIENT or other trained personnel under **medical supervision** ~~of medical expertise~~.

This document includes all ME EQUIPMENT that is intended to deliver a HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION treatment to a PATIENT ~~suffering from kidney failure~~, **independent of the treatment duration and location**.

If applicable, this document applies to the relevant parts of ME EQUIPMENT intended for other extracorporeal blood purification treatments.

The particular requirements in this document do not apply to:

- EXTRACORPOREAL CIRCUITS (see ISO 8637-2, [12]²);
- DIALYSERS (see ISO 8637-1, [11]);
- DIALYSIS FLUID CONCENTRATES (see ISO 23500-4, [18]);
- DIALYSIS WATER ~~treatment equipment~~ supply systems (see ISO 23500-2, [16]);
- CENTRAL DELIVERY SYSTEMS for DIALYSIS FLUID CONCENTRATES (see ISO 23500-4, [18]), described as systems for bulk mixing concentrate at a dialysis facility;
- equipment used to perform PERITONEAL DIALYSIS (see IEC 60601-2-39, [8]).

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

² Numbers in square brackets refer to the Bibliography.

~~If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.~~

~~HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of IEC 60601-1.~~

~~NOTE See also 4.2 of IEC 60601-1:2005.~~

201.1.2 Object

Replacement:

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for HAEMODIALYSIS EQUIPMENT.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2:2014, IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, IEC 60601-1-10:2007 and IEC 60601-1-10:2007/AMD1:2013 and IEC 60601-1-11:2015 apply as modified in Clauses 202, 208, 210 and 211 ~~respectively~~. IEC 60601-1-3 and IEC 60601-1-12 do not apply. IEC 60601-1-9:2007 and IEC 60601-1-9:2007/AMD1:2013 does not apply as noted in Clause 209. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"*Addition*" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"*Amendment*" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.439-147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, for example 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the bibliography.

Clause 2 of the general standard applies, except as follows:

Amendment Replacement:

IEC 60601-1-2:~~2007~~ 2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic-compatibility disturbances – Requirements and tests*

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*

IEC 60601-1-6:2010/AMD1:2013

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*

IEC 60601-1-8:2006/AMD1:2012

Addition:

IEC 60601-1-10:2007, *Medical electrical equipment – Part 1-10: General requirements for basic safety and essential performance – Collateral Standard: Requirements for the development of physiologic closed-loop controllers*

IEC 60601-1-10:2007/AMD1:2013

IEC 60601-1-11:~~2010~~ 2015, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 61672-1, *Electroacoustics – Sound level meters – Part 1: Specifications*

~~IEC 62366:2007, *Medical devices – Application of usability engineering to medical devices*~~

~~ISO 594-2, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment – Part 2: Lock fittings*~~

ISO 3744, *Acoustics – Determination of sound power levels and sound energy levels of noise sources using sound pressure – Engineering method in an essentially free field over a reflecting plane*

~~ISO 8638, *Cardiovascular implants and artificial organs – Extracorporeal blood circuit for haemodialysers, haemodiafilters and haemofilters*~~

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-16: Particular requirements for the basic safety and essential performance
of haemodialysis, haemodiafiltration and haemofiltration equipment**

**Appareils électromédicaux –
Partie 2-16: Exigences particulières pour la sécurité de base et les performances
essentielles des appareils d'hémodialyse, d'hémodiafiltration et d'hémofiltration**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards.....	8
201.2 Normative references	10
201.3 Terms and definitions	11
201.4 General requirements	14
201.5 General requirements for testing ME EQUIPMENT	17
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	17
201.7 ME EQUIPMENT identification, marking and documents	17
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	22
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	23
201.10 Protection against unwanted and excessive radiation HAZARDS	23
201.11 Protection against excessive temperatures and other HAZARDS	23
201.12 * Accuracy of controls and instruments and protection against hazardous outputs	24
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	34
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	34
201.15 Construction of ME EQUIPMENT.....	35
201.16 * ME SYSTEMS	35
201.17 ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS	36
202 Electromagnetic disturbances – Requirements and tests	36
208 General requirements, tests and guidance for ALARM SYSTEMS in MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS	37
209 Requirements for environmentally conscious design	39
210 Requirements for the development of PHYSIOLOGIC CLOSED-LOOP CONTROLLERS	39
211 * Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the HOME HEALTHCARE ENVIRONMENT.....	39
Annexes	41
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures.....	42
Annex AA (informative) Particular guidance and rationale	43
Annex BB (informative) Examples of HAZARDS, foreseeable sequences of events, and HAZARDOUS SITUATIONS in HAEMODIALYSIS EQUIPMENT	62
Bibliography.....	71
Index of defined terms used in this particular standard.....	74
Figure 201.101 – Continuous air infusion test setup with example dimensions	30
Figure AA.1 – Example of a HAEMODIALYSIS ME SYSTEM	58
Table 201.101 – ESSENTIAL PERFORMANCE requirements	14

Table AA.1 – Example of ALARM CONDITION priorities according to 6.1.2
of IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, adapted for
HAEMODIALYSIS EQUIPMENT needs..... 60

Table BB.1 – HAZARDOUS SITUATION list following ISO 14971:2007, Annex E 62

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-16: Particular requirements for the basic safety and essential performance of haemodialysis, haemodiafiltration and haemofiltration equipment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-16 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This fifth edition cancels and replaces the fourth edition of IEC 60601-2-16 published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) update of references to IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, of references and requirements to IEC 60601-1-2:2014, of references to IEC 60601-1-6:2010 and IEC 60601-1-6:2010/AMD1:2013, of references and requirements to IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, of references to

IEC 60601-1-9:2007 and IEC 60601-1-9:2007/AMD1:2013, of references to IEC 60601-1-10:2007 and IEC 60601-1-10:2007/AMD1:2013 and of references to IEC 60601-1-11:2015;

- b) widening of the scope;
- c) editorial improvements;
- d) addition of requirements for anticoagulant delivery means;
- e) other few small technical changes.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/1557/FDIS	62D/1585/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION EQUIPMENT.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-16: Particular requirements for the basic safety and essential performance of haemodialysis, haemodiafiltration and haemofiltration equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION EQUIPMENT, hereafter referred to as HAEMODIALYSIS EQUIPMENT.

This document does not take into consideration specific safety details of the DIALYSIS FLUID control system of HAEMODIALYSIS EQUIPMENT using regeneration of DIALYSIS FLUID or CENTRAL DELIVERY SYSTEMS for DIALYSIS FLUID. It does, however, take into consideration the specific safety requirements of such HAEMODIALYSIS EQUIPMENT concerning electrical safety and PATIENT safety.

This document specifies the minimum safety requirements for HAEMODIALYSIS EQUIPMENT. These HAEMODIALYSIS EQUIPMENT are intended for use either by medical staff or for use by the PATIENT or other trained personnel under medical supervision.

This document includes all ME EQUIPMENT that is intended to deliver a HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION treatment to a PATIENT, independent of the treatment duration and location.

If applicable, this document applies to the relevant parts of ME EQUIPMENT intended for other extracorporeal blood purification treatments.

The particular requirements in this document do not apply to:

- EXTRACORPOREAL CIRCUITS (see ISO 8637-2, [12]²);
- DIALYSERS (see ISO 8637-1, [11]);
- DIALYSIS FLUID CONCENTRATES (see ISO 23500-4, [18]);
- DIALYSIS WATER supply systems (see ISO 23500-2, [16]);
- CENTRAL DELIVERY SYSTEMS for DIALYSIS FLUID CONCENTRATES (see ISO 23500-4, [18]), described as systems for bulk mixing concentrate at a dialysis facility;
- equipment used to perform PERITONEAL DIALYSIS (see IEC 60601-2-39, [8]).

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

² Numbers in square brackets refer to the Bibliography.

201.1.2 Object

Replacement:

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for HAEMODIALYSIS EQUIPMENT.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2:2014, IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, IEC 60601-1-10:2007 and IEC 60601-1-10:2007/AMD1:2013 and IEC 60601-1-11:2015 apply as modified in Clauses 202, 208, 210 and 211. IEC 60601-1-3 and IEC 60601-1-12 do not apply. IEC 60601-1-9:2007 and IEC 60601-1-9:2007/AMD1:2013 does not apply as noted in Clause 209. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"*Addition*" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"*Amendment*" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are

numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, for example 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the bibliography.

Clause 2 of the general standard applies, except as follows:

Replacement:

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*
IEC 60601-1-6:2010/AMD1:2013

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*
IEC 60601-1-8:2006/AMD1:2012

Addition:

IEC 60601-1-10:2007, *Medical electrical equipment – Part 1-10: General requirements for basic safety and essential performance – Collateral Standard: Requirements for the development of physiologic closed-loop controllers*
IEC 60601-1-10:2007/AMD1:2013

IEC 60601-1-11:2015, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 61672-1, *Electroacoustics – Sound level meters – Part 1: Specifications*

ISO 3744, *Acoustics – Determination of sound power levels and sound energy levels of noise sources using sound pressure – Engineering method in an essentially free field over a reflecting plane*

SOMMAIRE

AVANT-PROPOS.....	80
INTRODUCTION.....	83
201.1 Domaine d'application, objet et normes connexes.....	84
201.2 Références normatives.....	86
201.3 Termes et définitions.....	87
201.4 Exigences générales.....	90
201.5 Exigences générales relatives aux essais des APPAREILS EM.....	94
201.6 Classification des APPAREILS EM et des SYSTEMES EM.....	94
201.7 Identification, marquage et documentation des APPAREILS EM.....	94
201.8 Protection contre les DANGERS d'origine électrique provenant des APPAREILS EM.....	99
201.9 Protection contre les DANGERS MECANQUES des APPAREILS EM et des SYSTEMES EM.....	100
201.10 Protection contre les DANGERS dus aux rayonnements involontaires ou excessifs.....	100
201.11 Protection contre les températures excessives et les autres DANGERS.....	100
201.12 * Précision des commandes, des instruments et protection contre les caractéristiques de sortie présentant des risques.....	101
201.13 SITUATIONS DANGEREUSES et conditions de défaut POUR LES APPAREILS EM.....	112
201.14 SYSTEMES ELECTROMEDICAUX PROGRAMMABLES (SEMP).....	112
201.15 Construction de l'APPAREIL EM.....	113
201.16 * SYSTEMES EM.....	114
201.17 COMPATIBILITE ELECTROMAGNETIQUE des APPAREILS et des SYSTEMES EM.....	114
202 Perturbations électromagnétiques – Exigences et essais.....	115
208 Exigences générales, essais et guide pour les SYSTEMES D'ALARME des APPAREILS et des SYSTEMES ELECTROMEDICAUX.....	115
209 Exigences pour une conception écoresponsable.....	117
210 Exigences pour le développement des REGULATEURS PHYSIOLOGIQUES EN BOUCLE FERMEE.....	117
211 * Exigences pour les APPAREILS ELECTROMEDICAUX et les SYSTEMES ELECTROMEDICAUX utilisés dans l'ENVIRONNEMENT DES SOINS A DOMICILE.....	118
Annexes.....	119
Annexe G (normative) Protection contre les DANGERS d'inflammation des mélanges anesthésiques inflammables.....	120
Annexe AA (informative) Guide particulier et justifications.....	121
Annexe BB (informative) Exemples de DANGERS, de séquences d'événements prévisibles et de SITUATIONS DANGEREUSES dans les APPAREILS D'HEMODIALYSE.....	142
Bibliographie.....	151
Index des termes définis utilisés dans la présente norme particulière.....	154
 Figure 201.101 – Montage d'essai pour l'infusion d'air continue avec des exemples de dimensions.....	 108
Figure AA.1 – Exemple de SYSTEME EM pour l'HEMODIALYSE.....	138

Tableau 201.101 – Exigences de PERFORMANCES ESSENTIELLES.....	90
Tableau AA.1 – Exemple de priorités de CONDITIONS D'ALARME conformément à 6.1.2 de l'IEC 60601-1-8:2006 et à l'IEC 60601-1-8:2006/AMD1:2012, adaptées aux besoins des APPAREILS D'HEMODIALYSE	140
Tableau BB.1 – Liste de SITUATIONS DANGEREUSES suivant l'Annexe E de l'ISO 14971:2007	142

COMMISSION ÉLECTROTECHNIQUE INTERNATIONALE

APPAREILS ÉLECTROMÉDICAUX –

Partie 2-16: Exigences particulières pour la sécurité de base et les performances essentielles des appareils d'hémodialyse, d'hémodiafiltration et d'hémofiltration

AVANT-PROPOS

- 1) La Commission Electrotechnique Internationale (IEC) est une organisation mondiale de normalisation composée de l'ensemble des comités électrotechniques nationaux (Comités nationaux de l'IEC). L'IEC a pour objet de favoriser la coopération internationale pour toutes les questions de normalisation dans les domaines de l'électricité et de l'électronique. À cet effet, l'IEC – entre autres activités – publie des Normes internationales, des Spécifications techniques, des Rapports techniques, des Spécifications accessibles au public (PAS) et des Guides (ci-après dénommés "Publication(s) de l'IEC"). Leur élaboration est confiée à des comités d'études, aux travaux desquels tout Comité national intéressé par le sujet traité peut participer. Les organisations internationales, gouvernementales et non gouvernementales, en liaison avec l'IEC, participent également aux travaux. L'IEC collabore étroitement avec l'Organisation Internationale de Normalisation (ISO), selon des conditions fixées par accord entre les deux organisations.
- 2) Les décisions ou accords officiels de l'IEC concernant les questions techniques représentent, dans la mesure du possible, un accord international sur les sujets étudiés, étant donné que les Comités nationaux de l'IEC intéressés sont représentés dans chaque comité d'études.
- 3) Les Publications de l'IEC se présentent sous la forme de recommandations internationales et sont agréées comme telles par les Comités nationaux de l'IEC. Tous les efforts raisonnables sont entrepris afin que l'IEC s'assure de l'exactitude du contenu technique de ses publications; l'IEC ne peut pas être tenue responsable de l'éventuelle mauvaise utilisation ou interprétation qui en est faite par un quelconque utilisateur final.
- 4) Dans le but d'encourager l'uniformité internationale, les Comités nationaux de l'IEC s'engagent, dans toute la mesure possible, à appliquer de façon transparente les Publications de l'IEC dans leurs publications nationales et régionales. Toutes divergences entre toutes Publications de l'IEC et toutes publications nationales ou régionales correspondantes doivent être indiquées en termes clairs dans ces dernières.
- 5) L'IEC elle-même ne fournit aucune attestation de conformité. L'IEC n'est responsable d'aucun des services effectués par les organismes de certification indépendants. L'IEC n'est responsable d'aucun des services effectués par les organismes de certification indépendants.
- 6) Tous les utilisateurs doivent assurer qu'ils sont en possession de la dernière édition de cette publication.
- 7) Aucune responsabilité ne doit être imputée à l'IEC, à ses administrateurs, employés, auxiliaires ou mandataires, y compris ses experts particuliers et les membres de ses comités d'études et des Comités nationaux de l'IEC, pour tout préjudice causé en cas de dommages corporels et matériels, ou de tout autre dommage de quelque nature que ce soit, directe ou indirecte, ou pour supporter les coûts (y compris les frais de justice) et les dépenses découlant de la publication ou de l'utilisation de cette Publication de l'IEC ou de toute autre Publication de l'IEC, ou au crédit qui lui est accordé.
- 8) L'attention est attirée sur les références normatives citées dans cette publication. L'utilisation de publications référencées est obligatoire pour une application correcte de la présente publication.
- 9) L'attention est attirée sur le fait que certains des éléments de la présente Publication de l'IEC peuvent faire l'objet de droits de brevet. L'IEC ne saurait être tenue pour responsable de ne pas avoir identifié de tels droits de brevets et de ne pas avoir signalé leur existence.

La Norme internationale IEC 60601-2-16 a été établie par le sous-comité 62D: Appareils électromédicaux, du comité d'études 62 de l'IEC: Équipements électriques dans la pratique médicale.

Cette cinquième édition annule et remplace la quatrième édition de l'IEC 60601-2-16 parue en 2012. Cette édition constitue une révision technique.

Cette édition inclut les modifications techniques majeures suivantes par rapport à l'édition précédente:

- a) actualisation des références à l'IEC 60601-1:2005 et l'IEC 60601-1:2005/AMD1:2012, des références et des exigences à l'IEC 60601-1-2:2014, des références à l'IEC 60601-1-6:2010 et l'IEC 60601-1-6:2010/AMD1:2013, des références et des exigences à

l'IEC 60601-1-8:2006 et l'IEC 60601-1-8:2006/AMD1:2012, des références à l'IEC 60601-1-9:2007 et l'IEC 60601-1-9:2007/AMD1:2013, des références à l'IEC 60601-1-10:2007 et l'IEC 60601-1-10:2007/AMD1:2013 ainsi que des références à l'IEC 60601-1-11:2015;

- b) élargissement du domaine d'application;
- c) améliorations d'ordre rédactionnel;
- d) ajout d'exigences concernant les dispositifs de transmission d'anticoagulant;
- e) quelques autres modifications techniques limitées.

Le texte de cette norme particulière est issu des documents suivants:

FDIS	Rapport de vote
62D/1557/FDIS	62D/1585/RVD

Le rapport de vote indiqué dans le tableau ci-dessus donne toute information sur le vote ayant abouti à l'approbation de cette norme particulière.

Cette publication a été rédigée selon les Directives ISO/IEC, Partie 2.

Dans le présent document, les caractères d'imprimerie suivants sont utilisés:

- exigences et définitions: caractères romains;
- *modalités d'essais: caractères italiques;*
- indications de nature informative apparaissant hors des tableaux, comme les notes, les exemples et les références: petits caractères. Le texte normatif à l'intérieur des tableaux est également en petits caractères;
- TERMES DEFINIS A L'ARTICLE 3 DE LA NORME GENERALE, DE LA PRESENTE NORME PARTICULIERE OU COMME NOTES: PETITES MAJUSCULES.

Concernant la structure du présent document, le terme

- "article" désigne l'une des dix-sept divisions numérotées dans la table des matières, avec toutes ses subdivisions (par exemple, l'Article 7 inclut les paragraphes 7.1, 7.2, etc.);
- "paragraphe" désigne une subdivision numérotée d'un article (par exemple 7.1, 7.2 et 7.2.1 sont tous des paragraphes appartenant à l'Article 7).

Dans le présent document, les références à des articles sont précédées du mot "Article" suivi du numéro de l'article concerné. Dans la présente norme particulière, les références aux paragraphes utilisent uniquement le numéro du paragraphe concerné.

Dans le présent document, la conjonction "ou" est utilisée avec la valeur d'un "ou inclusif", ainsi un énoncé est vrai si une combinaison des conditions, quelle qu'elle soit, est vraie.

Les formes verbales utilisées dans le présent document sont conformes à l'usage donné à l'Article 7 des Directives ISO/IEC, Partie 2. Pour les besoins du présent document:

- "devoir" mis au présent de l'indicatif signifie que la satisfaction à une exigence ou à un essai est obligatoire pour la conformité au présent document;
- "il convient" signifie que la satisfaction à une exigence ou à un essai est recommandée, mais n'est pas obligatoire pour la conformité au présent document;
- "pouvoir" mis au présent de l'indicatif est utilisé pour décrire un moyen admissible pour satisfaire à une exigence ou à un essai.

Lorsqu'un astérisque (*) est utilisé comme premier caractère devant un titre ou au début d'un titre d'alinéa ou de tableau, il indique l'existence d'un guide ou d'une justification à consulter à l'Annexe AA.

Une liste de toutes les parties de la série IEC 60601, publiées sous le titre général *Appareils électromédicaux*, peut être consultée sur le site web de l'IEC.

Le comité a décidé que le contenu de cette publication ne sera pas modifié avant la date de stabilité indiquée sur le site web de l'IEC sous "<http://webstore.iec.ch>" dans les données relatives à la publication recherchée. À cette date, la publication sera

- reconduite,
- supprimée,
- remplacée par une édition révisée, ou
- amendée.

IMPORTANT – Le logo "*colour inside*" qui se trouve sur la page de couverture de cette publication indique qu'elle contient des couleurs qui sont considérées comme utiles à une bonne compréhension de son contenu. Les utilisateurs devraient, par conséquent, imprimer cette publication en utilisant une imprimante couleur.

NOTE L'attention des utilisateurs du présent document est attirée sur le fait que les fabricants d'appareils et les organismes d'essai peuvent avoir besoin d'une période transitoire après la publication d'une nouvelle publication IEC, ou d'une publication amendée ou révisée, pour fabriquer des produits conformes aux nouvelles exigences et pour adapter leurs équipements aux nouveaux essais ou aux essais révisés. Le comité recommande que le contenu de cette publication soit entériné au niveau national au plus tôt 3 ans après la date de publication.

INTRODUCTION

Les exigences minimales de sécurité spécifiées dans la présente norme particulière sont considérées comme fournissant un degré pratique de sécurité pour le fonctionnement des APPAREILS D'HEMODIALYSE, D'HEMODIAFILTRATION et D'HEMOPILTRATION.

APPAREILS ÉLECTROMÉDICAUX –

Partie 2-16: Exigences particulières pour la sécurité de base et les performances essentielles des appareils d'hémodialyse, d'hémodiafiltration et d'hémofiltration

201.1 Domaine d'application, objet et normes connexes

L'Article 1 de la norme générale¹ s'applique, avec les exceptions suivantes:

201.1.1 * Domaine d'application

Remplacement:

La présente partie de l'IEC 60601 s'applique à la SECURITE DE BASE et aux PERFORMANCES ESSENTIELLES des APPAREILS D'HEMODIALYSE, D'HEMODIAFILTRATION et D'HEMOfILTRATION, désignés ci-après sous le terme d'APPAREILS D'HEMODIALYSE.

Le présent document ne prend pas en considération les informations détaillées de sécurité spécifiques du système de contrôle du LIQUIDE DE DIALYSE de l'APPAREIL D'HEMODIALYSE utilisant la régénération du LIQUIDE DE DIALYSE ou les SYSTEMES DE TRANSMISSION CENTRALISES pour le LIQUIDE DE DIALYSE. Toutefois, il prend en considération les exigences de sécurité spécifiques de l'APPAREIL D'HEMODIALYSE concernant la sécurité électrique et la sécurité du PATIENT.

Le présent document spécifie les exigences minimales de sécurité relatives aux APPAREILS D'HEMODIALYSE. Ces APPAREILS D'HEMODIALYSE sont destinés à être utilisés soit par le personnel médical, soit par le PATIENT, soit par d'autres personnes formées, sous le contrôle d'un personnel ayant une compétence médicale.

Le présent document s'applique à tous les APPAREILS EM destinés à fournir un traitement d'HEMODIALYSE, d'HEMODIAFILTRATION et d'HEMOfILTRATION à un PATIENT, indépendamment de la durée et du lieu de traitement.

Le cas échéant, le présent document s'applique aux parties correspondantes des APPAREILS EM destinés à d'autres traitements extracorporels de purification du sang.

Les exigences particulières du présent document ne s'appliquent pas aux:

- CIRCUITS EXTRACORPORELS (voir ISO 8637-2 [12]²);
- DIALYSEURS (voir ISO 8637-1 [11]);
- CONCENTRES DE LIQUIDE DE DIALYSE (voir ISO 23500-4 [18]);
- systèmes d'approvisionnement en EAU DE DIALYSE (voir ISO 23500-2 [16]);
- SYSTEMES DE TRANSMISSION CENTRALISES pour LES CONCENTRES DE LIQUIDE DE DIALYSE (voir ISO 23500-4 [18]), décrits comme systèmes de mélange de concentré en vrac dans un centre de dialyse;

¹ La norme générale est l'IEC 60601-1:2005 et l'IEC 60601-1:2005/AMD1:2012, *Appareils électromédicaux – Partie 1: Exigences générales pour la sécurité de base et les performances essentielles*.

² Les chiffres entre crochets se réfèrent à la Bibliographie.

– appareils de DIALYSE PERITONEALE (voir IEC 60601-2-39 [8]).

201.1.2 Objet

Remplacement:

L'objet de la présente norme particulière est d'établir les exigences pour la SECURITE DE BASE et les PERFORMANCES ESSENTIELLES des APPAREILS D'HEMODIALYSE.

201.1.3 Normes collatérales

Addition:

La présente norme particulière fait référence aux normes collatérales applicables énumérées à l'Article 2 de la norme générale et à l'Article 201.2 de la présente norme particulière.

L'IEC 60601-1-2:2014, l'IEC 60601-1-8:2006 et l'IEC 60601-1-8:2006/AMD1:2012, l'IEC 60601-1-10:2007 et l'IEC 60601-1-10:2007/AMD1:2013, ainsi que l'IEC 60601-1-11:2015 s'appliquent telles que modifiées dans les Articles 202, 208, 210 et 211. L'IEC 60601-1-3 et l'IEC 60601-1-12 ne s'appliquent pas. L'IEC 60601-1-9:2007 et l'IEC 60601-1-9:2007/AMD1:2013 ne s'appliquent pas comme indiqué à l'Article 209. Toutes les autres normes collatérales publiées de la série IEC 60601-1 s'appliquent telles qu'elles sont publiées.

201.1.4 Normes particulières

Remplacement:

Dans la série IEC 60601, des normes particulières peuvent modifier, remplacer ou supprimer des exigences contenues dans la norme générale et dans les normes collatérales en fonction de ce qui est approprié à l'APPAREIL EM à l'étude, et elles peuvent ajouter d'autres exigences de SECURITE DE BASE et de PERFORMANCES ESSENTIELLES.

Une exigence d'une norme particulière prévaut sur l'exigence de la norme générale.

Par souci de concision, dans la présente norme particulière, le terme "norme générale" désigne les normes IEC 60601-1:2005 et IEC 60601-1:2005/AMD1:2012. Les normes collatérales sont désignées par leur numéro de document.

La numérotation des articles et paragraphes de la présente norme particulière correspond à celle de la norme générale avec le préfixe "201" (par exemple 201.1 dans le présent document aborde le contenu de l'Article 1 de la norme générale) ou de la norme collatérale applicable avec le préfixe "20x" où x est (sont) le (les) dernier(s) chiffre(s) du numéro de document de la norme collatérale (par exemple 202.4 dans la présente norme particulière aborde le contenu de l'Article 4 de la norme collatérale IEC 60601-1-2, 203.4 dans la présente norme particulière aborde le contenu de l'Article 4 de la norme collatérale IEC 60601-1-3, etc.). Les modifications apportées au texte de la norme générale sont précisées en utilisant les termes suivants:

"Remplacement" signifie que l'article ou le paragraphe de la norme générale ou de la norme collatérale applicable est remplacé complètement par le texte de la présente norme particulière.

"Addition" signifie que le texte de la présente norme particulière vient s'ajouter aux exigences de la norme générale ou de la norme collatérale applicable.

"*Amendement*" signifie que l'article ou le paragraphe de la norme générale ou de la norme collatérale applicable est modifié comme indiqué par le texte de la présente norme particulière.

Les paragraphes, les figures ou les tableaux qui sont ajoutés à ceux de la norme générale sont numérotés à partir de 201.101. Toutefois, en raison du fait que les définitions dans la norme générale sont numérotées de 3.1 à 3.147, les définitions complémentaires dans le présent document sont numérotées à partir de 201.3.201. Les annexes supplémentaires sont notées AA, BB, etc., et les points complémentaires aa), bb), etc.

Les paragraphes, les figures ou les tableaux qui sont ajoutés à ceux d'une norme collatérale sont numérotés à partir de 20x, où "x" est le chiffre de la norme collatérale, par exemple 202 pour l'IEC 60601-1-2, 203 pour l'IEC 60601-1-3, etc.

L'expression "le présent document" est utilisée pour se référer à la norme générale, à toutes les normes collatérales applicables et à la présente norme particulière, considérées ensemble.

Lorsque la présente norme particulière ne comprend pas d'article ou de paragraphe correspondant, l'article ou le paragraphe de la norme générale ou de la norme collatérale applicable, qui peut être sans objet, s'applique sans modification; lorsqu'il est demandé qu'une partie quelconque de la norme générale ou de la norme collatérale applicable, bien que pertinente, ne s'applique pas, cela est expressément mentionné dans la présente norme particulière.

201.2 Références normatives

NOTE Les références informatives sont énumérées dans la bibliographie.

L'Article 2 de la norme générale s'applique, avec les exceptions suivantes:

Remplacement:

IEC 60601-1-2:2014, *Appareils électromédicaux – Partie 1-2: Exigences générales pour la sécurité de base et les performances essentielles – Norme collatérale: Perturbations électromagnétiques – Exigences et essais*

IEC 60601-1-6:2010, *Appareils électromédicaux – Partie 1-6: Exigences générales pour la sécurité de base et les performances essentielles – Norme collatérale: Aptitude à l'utilisation*
IEC 60601-1-6:2010/AMD1:2013

IEC 60601-1-8:2006, *Appareils électromédicaux – Partie 1-8: Exigences générales pour la sécurité de base et les performances essentielles – Norme collatérale: Exigences générales, essais et guide pour les systèmes d'alarme des appareils et des systèmes électromédicaux*
IEC 60601-1-8:2006/AMD1:2012

Addition:

IEC 60601-1-10:2007, *Appareils électromédicaux – Partie 1-10: Exigences générales pour la sécurité de base et les performances essentielles – Norme collatérale: Exigences pour le développement des régulateurs physiologiques en boucle fermée*
IEC 60601-1-10:2007/AMD1:2013

IEC 60601-1-11:2015, *Appareils électromédicaux – Partie 1-11: Exigences générales pour la sécurité de base et les performances essentielles – Norme Collatérale: Exigences pour les appareils électromédicaux et les systèmes électromédicaux utilisés dans l'environnement des soins à domicile*

IEC 61672-1, *Electroacoustique – Sonomètres – Partie 1: Spécifications*

ISO 3744, *Acoustique – Détermination des niveaux de puissance acoustique et des niveaux d'énergie acoustique émis par les sources de bruit à partir de la pression acoustique – Méthode d'expertise pour des conditions approchant celles du champ libre sur plan réfléchissant*