

INTERNATIONAL STANDARD

IEC 60601-2-17

Second edition
2004-01

Medical electrical equipment –

Part 2-17: Particular requirements for the safety of automatically-controlled brachytherapy afterloading equipment

© IEC 2004 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembé, PO Box 131, CH-1211 Geneva 20, Switzerland
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: inmail@iec.ch Web: www.iec.ch



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE

V

For price, see current catalogue

CONTENTS

FOREWORD	3
INTRODUCTION	5
SECTION ONE – GENERAL	
1 Scope and object	6
2 Terminology and definitions	8
4 General requirements for tests	10
5 Classification	10
6 Identification, marking and documents	11
SECTION TWO – ENVIRONMENTAL CONDITIONS	
10 Environmental conditions	15
SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS.	
SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
29 X-radiation	15
30 Alpha, beta, gamma, NEUTRON RADIATION and other particle radiation	15
SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES	
SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS	
SECTION EIGHT – ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT	
50 Accuracy of operating data	25
SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS	
52 Abnormal operation and fault conditions	28
SECTION TEN – CONSTRUCTIONAL REQUIREMENTS	
Appendix L (normative) References – Publications mentioned in this standard	29
Bibliography	30
Index of defined terms	31

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-17: Particular requirements for the safety of
automatically-controlled brachytherapy afterloading equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-17 has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 1989, and its Amendment 1 (1996). Consideration has been given to new IEC standards, amendments to existing IEC standards, developments in technology and clinical usage, and various hazards encountered and envisaged since the preparation of the first edition.

The text of this standard is based on the following documents:

FDIS	Report on voting
62C/363/FDIS	62C/366/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this Particular Standard, the following print types are used:

- requirements, compliance with which can be tested, and definitions: roman type;
- explanations, advice, notes, general statements, exceptions and references: small roman type;
- *test specifications: italic type;*
- TERMS USED THROUGHOUT THIS PARTICULAR STANDARD WHICH ARE DEFINED IN CLAUSE 2 OR IN THE GENERAL STANDARD IEC 60601-1, ITS COLLATERAL OR PART 2 STANDARDS, OR IN IEC 60788: SMALL CAPITALS.

NOTE Attention is drawn to the existence, in some countries, of legislation containing requirements for:

- IONIZING RADIATION safety which may not align with the provisions of this Particular Standard, and
- maintenance, quality assurance and other related subjects, which are not covered by this Standard.

The committee has decided that the contents of this publication will remain unchanged until 2008. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

INTRODUCTION

The use of AFTERLOADING EQUIPMENT for BRACHYTHERAPY purposes may expose PATIENTS to danger if the EQUIPMENT fails to deliver the required dose to the PATIENT, or if the EQUIPMENT design does not satisfy standards of electrical and mechanical safety. The EQUIPMENT may also cause danger to persons in the vicinity if the EQUIPMENT itself fails to contain the RADIOACTIVE SOURCE(S) adequately within the STORAGE CONTAINER(S) and/or if there are inadequacies in the design of the TREATMENT ROOM.

This Particular Standard establishes requirements to be complied with by MANUFACTURERS in the design and construction of AFTERLOADING EQUIPMENT for use in temporary BRACHYTHERAPY procedures; it does not attempt to define their optimum performance requirements. Its purpose is to identify those features of design that are regarded, at the present time, as essential for the safe operation of such EQUIPMENT. It places limits on the degradation of EQUIPMENT performance beyond which it can be presumed that a fault condition exists and where an INTERLOCK then operates to return the RADIOACTIVE SOURCE(S) to the STORAGE CONTAINER(S) and afterwards to prevent continued operation of the EQUIPMENT.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-17: Particular requirements for the safety of automatically-controlled brachytherapy afterloading equipment

SECTION ONE – GENERAL

The clauses and subclauses of this section of the General Standard apply except as follows:

1 Scope and object

This clause of the General Standard applies except as follows:

1.1 Scope

Addition:

1.1.101 This Particular Standard specifies requirements for the safety of automatically-controlled EQUIPMENT for BRACHYTHERAPY of patients using AFTERLOADING techniques.

1.1.102 This Standard specifies requirements for automatically-controlled AFTERLOADING EQUIPMENT

- which contains and uses only beta, gamma, and NEUTRON-emitting SEALED RADIOACTIVE SOURCES,
- which automatically drives the sealed radioactive source(s) from a storage container to a treatment position inside the source applicator(s) and returns the source(s) to the storage container,
- which is designed for connection to a PATIENT, and
- with which movements of the RADIOACTIVE SOURCE(S) are carried out automatically by the EQUIPMENT according to a prescribed programme using a powered mechanism whose changes are controlled by the CONTROLLING TIMER(S) and TIMING DEVICES that are either PROGRAMMABLE ELECTRONIC SUB-SYSTEMS (PESS) (computer or microprocessors) or non-programmable systems.

1.1.103 This Standard specifies requirements for EQUIPMENT intended to be

- used under the supervision of QUALIFIED PERSONS;
- maintained at predetermined intervals;
- subject to regular checks by the user.

This Standard does not specify requirements for SEALED RADIOACTIVE SOURCES used with the EQUIPMENT. Such requirements are specified in other standards (see 6.8.3).

1.1.104 The requirements of this Standard are based on the assumptions that:

- a TREATMENT PLAN is available that prescribes appropriate values of the TREATMENT PARAMETERS, and
- the SOURCE STRENGTH(S) of the RADIOACTIVE SOURCE(S) used by the EQUIPMENT is (are) known.

This Standard includes requirements intended to ensure that the prescribed values of the TREATMENT PARAMETERS can be achieved by the EQUIPMENT, in particular that:

- the selected RADIOACTIVE SOURCE(S) is (are) positioned or moved within the SOURCE APPLICATOR in the selected configuration relative to the SOURCE APPLICATOR;
- IRRADIATION is performed by the selected RADIOACTIVE SOURCE configuration for the selected duration;
- IRRADIATION is performed by the EQUIPMENT without causing unnecessary risk to the OPERATOR or other persons in the immediate surroundings.

1.2 Object

Replacement:

The object of this Particular Standard is to establish particular requirements for safety of automatically-controlled BRACHYTHERAPY AFTERLOADING EQUIPMENT and the compliance test specifications. It presents the general functional requirements of the demand for safety rather than particular technological means of implementation.

1.3 Particular Standards

ADDITION:

1.3.101 Relationship to the General Standard

This Particular Standard is to be read in conjunction with IEC 60601-1:1988, *Medical electrical equipment – Part 1: General requirements for safety*, with its Amendments 1 (1991) and 2 (1995) – hereinafter referred to as the General Standard – which it amends and supplements. As in the General Standard, the requirements are followed by compliance tests.

The numbering of sections, clauses and subclauses of this Particular Standard corresponds to that of the General Standard. The changes to the text of the General and Collateral Standards are specified by the use of the following words:

"Replacement" means that the clause or subclause of the General Standard is replaced completely by the text of this Particular Standard;

"Addition" means that the text of this Particular Standard is additional to the requirements of the General Standard;

"Amendment" means that the clause or subclause of the General Standard is amended as indicated by the text of this Particular Standard.

Subclauses, figures or tables which are additional to those of the General Standard are numbered starting from 101: additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

The term "this Standard " is used to make reference to the General Standard and this Particular Standard taken together.

Where there is no corresponding section, clause or subclause in this Particular Standard, the section, clause or subclause of the General Standard, although possibly not relevant, applies without modification.

Where it is intended that any part of the General Standard, although possibly relevant, is not to be applied, a statement to that effect is given in this Particular Standard.

The requirements of this Particular Standard take priority over those of the General Standard.

1.3.102 Relationship to other standards and documents

NOTE See Appendix L for normative references.

1.5 Collateral Standards

Addition:

1.5.101 IEC 60601-1-1

All clauses and subclauses of Collateral Standard IEC 60601-1-1 apply.

1.5.102 IEC 60601-1-2

All clauses and subclauses of Collateral Standard IEC 60601-1-2 apply.

1.5.103 IEC 60601-1-3

Collateral Standard IEC 60601-1-3 does not apply.

1.5.104 IEC 60601-1-4

All clauses and subclauses of Collateral Standard IEC 60601-1-4 apply.

Appendix L (normative)

References – Publications mentioned in this standard

This appendix of the General Standard applies, except as follows:

Addition:

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

Addition:

IEC 60788:1984, *Medical radiology – Terminology*

IEC 61217:1996, *Radiotherapy equipment —Coordinates, movements and scales* ¹⁾
Amendment 1 (2000)

IEC 62083:2000, *Medical electrical equipment – Requirements for the safety of radiotherapy treatment planning systems*

Replacement:

IEC 60601-1:1988, *Medical electrical equipment – Part 1 General requirements for safety*
Amendment 1 (1991)
Amendment 2 (1995)

IEC 60601-1-1:2000, *Medical electrical equipment – Part 1-1: General requirements for safety – Collateral standard: Safety requirements for medical electrical systems*

IEC 60601-1-2:2001, *Medical electrical equipment – Part 1-2: General requirements for safety –Collateral Standard: Electromagnetic compatibility – Requirements and tests*

IEC 60601-1-4:1996, *Medical electrical equipment – Part 1: General requirements for safety – 4. Collateral Standard: Programmable electrical medical systems* ²⁾
Amendment 1 (1999)

¹⁾ There exists a consolidated edition 1.1 (2000) that includes IEC 61217 (1996) and its Amendment 1 (2000).

²⁾ There exists a consolidated edition 1.1 (2000) that includes IEC 60601-1-4 (1996) and its Amendment 1 (1999).