

INTERNATIONAL STANDARD

IEC 60601-2-44

Edition 2.1

2002-11

Edition 2:2001 consolidée par l'amendement 1:2002
Edition 2:2001 consolidated with amendment 1:2002

Medical electrical equipment –

Part 2-44: Particular requirements for the safety of X-ray equipment for computed tomography

Appareils électromédicaux –

*Partie 2-44:
Règles particulières de sécurité pour les équipements
à rayonnement X de tomodensitométrie*

© IEC 2002 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembé, PO Box 131, CH-1211 Geneva 20, Switzerland
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: inmail@iec.ch Web: www.iec.ch



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE

W

For price, see current catalogue

CONTENTS

FOREWORD 4

SECTION 1: GENERAL

1 Scope and object 6
1.1 Scope 6
1.2 Object 6
1.3 Particular Standards 7
2 Terminology and definitions 8
3 General requirements 11
5 Classification 11
6 Identification, marking and documents 11
6.1 Marking on the outside of EQUIPMENT or EQUIPMENT parts 11
6.7 Indicator lights and push-buttons 12
6.8 ACCOMPANYING DOCUMENTS 12

SECTION 2: ENVIRONMENTAL CONDITIONS

10 Environmental conditions 14

SECTION 3: PROTECTION AGAINST ELECTRIC SHOCK HAZARDS

15 Limitation of voltage and/or energy 15
16 ENCLOSURES and PROTECTIVE COVERS 15
19 Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS 15
19.3 Allowable values 15
20 Dielectric strength 16
20.3 Values of test voltages 16
20.4 Tests 16

SECTION 4: PROTECTION AGAINST MECHANICAL HAZARDS

22 Moving parts 18
27 Pneumatic and hydraulic power 19

SECTION 5: PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION

29 X-RADIATION 20
36 Electromagnetic compatibility 28

SECTION 6: PROTECTION AGAINST HAZARDS OF IGNITION OF
FLAMMABLE ANAESTHETIC MIXTURES

SECTION 7: PROTECTION AGAINST EXCESSIVE TEMPERATURES AND
OTHER SAFETY HAZARDS

42	Excessive temperatures	29
----	------------------------------	----

SECTION 8: ACCURACY OF OPERATING DATA AND PROTECTION
AGAINST HAZARDOUS OUTPUT

50	Accuracy of operating data	29
	50.101 Accuracy of radiation output	29
	50.102 Accuracy of recorded examination data	29
51	Protection against hazardous output.....	30

SECTION 9: ABNORMAL OPERATION AND FAULT CONDITIONS;
ENVIRONMENTAL TESTS

SECTION 10: CONSTRUCTIONAL REQUIREMENTS

56	Components and general assembly	31
	56.7 Batteries	31
57	MAINS PARTS, components and layout	31
	57.10 CREEPAGE DISTANCES and AIR CLEARANCES	31

Tables

101	HALF-VALUE LAYERS in CT SCANNERS	26
-----	--	----

Figures

101	Coordinate system	9
-----	-------------------------	---

Annex AA (normative)	Terminology – Index of defined terms	32
----------------------	--	----

Annex BB (informative)	Choosing LOADING FACTORS for tests.....	35
------------------------	---	----

Bibliography.....		36
-------------------	--	----

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-44: Particular requirements for the safety of X-ray equipment for computed tomography

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-44 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This consolidated version of IEC 60601-2-44 is based on the second edition (2001) [documents 62B/426/FDIS and 62B/437/RVD] and its amendment 1 (2002) [documents 62B/472/FDIS and 62B/478/RVD].

It bears the edition number 2.1.

A vertical line in the margin shows where the base publication has been modified by amendment 1.

Annex AA forms an integral part of this standard.

Annex BB is for information only.

In this standard, the following print types are used:

- requirements, compliance with which can be tested and definitions: roman type;
- explanations, advice, notes, general statements and exceptions: smaller type;
- *test specifications and headings of subclauses: italic type;*
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD OR IN IEC 60788: SMALL CAPITALS.

The committee has decided that the contents of the base publication and its amendment 1 will remain unchanged until 2004. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-44: Particular requirements for the safety of X-ray equipment for computed tomography

SECTION 1: GENERAL

The clauses and subclauses of this section of the General Standard apply except as follows:

1 Scope and object

This clause of the General Standard applies except as follows:

1.1 Scope

Addition:

This Particular Standard applies to X-RAY EQUIPMENT for COMPUTED TOMOGRAPHY (CT SCANNERS).

It includes safety requirements for the X-RAY GENERATOR, and those where HIGH VOLTAGE GENERATORS are integrated with an X-RAY TUBE ASSEMBLY.

1.2 Object

Replacement:

The object of this standard is to establish particular requirements to ensure safety, and to specify methods for demonstrating compliance with those requirements, for CT SCANNERS.

NOTE 1 Requirements for reproducibility, linearity, constancy and accuracy are given because of their relationship to the quality and quantity of the IONIZING RADIATION produced and are confined to those considered necessary for safety.

NOTE 2 Both the levels for compliance and the tests prescribed to determine compliance reflect the fact that the safety of HIGH-VOLTAGE GENERATORS is not sensitive to small differences in levels of performance. The combinations of LOADING FACTORS specified for the tests are therefore limited in number but chosen from experience as being appropriate in most cases. It is considered important to standardize the choice of combinations of LOADING FACTORS so that comparison can be made between tests performed in different places on different occasions. However, combinations other than those specified could be of equal technical validity.

NOTE 3 The safety philosophy on which this standard is based is described in the introduction to the General Standard and in IEC 60513.

NOTE 4 Concerning RADIOLOGICAL PROTECTION it has been assumed in the preparation of this standard that MANUFACTURERS and USERS do accept the general principles of the ICRP as stated in ICRP 60, 1990, paragraph 112,¹⁾ namely:

"(a) No practice involving exposures to radiation should be adopted unless it produces sufficient benefit to the exposed individuals or to society to offset the radiation detriment it causes. (The justification of a practice.)

(b) In relation to any particular source within a practice, the magnitude of individual doses, the number of people exposed, and the likelihood of incurring exposures where these are not certain to be received should all be kept as low as reasonably achievable, economic and social factors being taken into account. This procedure should be constrained by restrictions on the doses to individuals (dose constraints), or the risks to individuals in the case of potential exposures (risk constraints), so as to limit the inequity likely to result from the inherent economic and social judgements. (The optimisation of protection.)

¹⁾ ICRP Publication 60: *Recommendations of the International Commission on Radiological Protection (Annals of the ICRP Vol. 21 No 1-3, 1990)*. Published by Pergamon Press.

(c) The exposure of individuals resulting from the combination of all the relevant practices should be subject to dose limits, or to some control of risk in the case of potential exposures. These are aimed at ensuring that no individual is exposed to radiation risks that are judged to be unacceptable from these practices in any normal circumstances. Not all sources are susceptible of control by action at the source and it is necessary to specify the sources to be included as relevant before selecting a dose limit. (Individual dose and risk limits.)"

NOTE 5 Most of the requirements on X-RAY EQUIPMENT and its sub-assemblies for protection against IONIZING RADIATION are given in the Collateral Standard IEC 60601-1-3.

This standard does, however, deal with some aspects of RADIOLOGICAL PROTECTION, mainly those that depend upon the supply, control and indication of electrical energy from the HIGH-VOLTAGE GENERATOR.

NOTE 6 It is recognized that many of the judgements necessary to follow the ICRP general principles have to be made by the USER and not by the MANUFACTURER of the EQUIPMENT.

1.3 Particular Standards

Addition:

This Particular Standard, hereinafter referred to as "this standard", amends and supplements a set of IEC publications, hereinafter referred to as "General Standard", consisting of IEC 60601-1:1988, *Medical electrical equipment – Part 1: General requirements for safety*, its amendments No. 1 (1991) and No. 2 (1995), and all Collateral Standards. The numbering of sections, clauses and subclauses of this standard corresponds to that of the General Standard. The changes to the text of the General Standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the General Standard is replaced completely by the text of this standard.

"Addition" means that the text of this standard is additional to the requirements of the General Standard.

"Amendment" means that the clause or subclause of the General Standard is amended as indicated by the text of this standard.

Subclauses or figures which are additional to those of the General Standard are numbered starting from 101, additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Where there is no corresponding section, clause or subclause in this standard, the section, clause or subclause of the General Standard applies without modification.

Where it is intended that any part of the General Standard, although possibly relevant, is not to be applied, a statement to that effect is given in this standard.

A requirement of this standard replacing or modifying requirements of the General Standard takes precedence over the original requirements concerned.

1.3.101 Related International Standards

IEC 60601-1-2:1993, *Medical electrical equipment – Part 1: General requirements for safety – 2. Collateral Standard: Electromagnetic compatibility – Requirements and tests*

IEC 60601-1-3:1994, *Medical electrical equipment – Part 1: General requirements for safety – 3. Collateral Standard: General requirements for radiation protection in diagnostic X-ray equipment*

IEC 60601-2-28:1993, *Medical electrical equipment – Part 2: Particular requirements for the safety of X-ray source assemblies and X-ray tube assemblies for medical diagnosis*

IEC 60664-1:1992, *Insulation coordination for equipment within low-voltage systems – Part 1: Principles, requirements and tests*

IEC 60788:1984, *Medical radiology – Terminology*

ISO 2092:1981, *Light metals and their alloys – Code of designation based on chemical symbols*