



IEC 61267

Edition 3.0 2026-03

INTERNATIONAL STANDARD

COMMENTED VERSION

**Medical diagnostic X-ray equipment - Radiation conditions for use in
the determination of characteristics**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	7
3 Terms and definitions	7
3.1 Terms defined in this document	7
3.2 Terms defined in other standards.....	8
4 Common aspects.....	16
4.1 X-ray radiation conditions	16
4.2 Test instrumentation	16
4.2.1 Maximum inherent filtration.....	16
4.2.2 Anode material	16
4.2.3 Percentage ripple of the X-ray tube voltage	16
4.2.4 X-ray tube voltage adjustment	16
4.2.5 X-ray tube voltage measuring device	17
4.2.6 Radiation detector	17
4.2.7 Measuring arrangements	17
5 RQR X-ray radiation conditions.....	17
5.1 Object.....	17
5.2 Characterization	17
5.3 Description	17
5.4 Additional filtration	18
5.5 Test equipment.....	18
5.5.1 Auxiliary filter	18
5.5.2 Attenuation layers.....	18
5.5.3 Diaphragm.....	19
5.5.4 Radiation detector	19
5.5.5 X-ray tube voltage measuring device	19
5.6 Generation and verification	19
5.6.1 Geometry	19
5.6.2 RQR X-ray radiation conditions.....	19
5.6.3 Series of RQR X-ray radiation conditions.....	19
6 RQA X-ray radiation conditions.....	20
6.1 Object.....	20
6.2 Characterization	20
6.3 Description	20
6.4 Added filter	21
6.5 Generation and verification	21
6.5.1 RQA X-ray radiation conditions.....	21
6.5.2 Alternative method.....	21
7 RQC X-ray radiation conditions.....	22
7.1 Object.....	22
7.2 Characterization	22
7.3 Description	22
7.4 Added filter	22
7.5 Generation and verification	23

8	RQT X-ray radiation conditions	23
8.1	Object.....	23
8.2	Characterization	23
8.3	Description	23
8.4	Added filter	24
8.5	Generation and verification	24
8.5.1	RQT X-ray radiation conditions	24
8.5.2	Alternative method.....	24
9	RQN X-ray radiation conditions.....	24
9.1	Object.....	24
9.2	Characterization	24
9.3	Description	25
9.4	Phantom	25
9.5	Diaphragms	25
9.6	Generation.....	25
10	RQB X-ray radiation conditions.....	26
10.1	Object.....	26
10.2	Characterization	26
10.3	Description	26
10.4	Phantom	26
10.5	Diaphragms	27
10.6	Generation.....	27
11	RQR-M X-ray radiation conditions	27
11.1	Object.....	27
11.2	Characterization	27
11.3	Description	27
11.4	Generation and verification	28
12	RQA-M X-ray radiation conditions.....	28
12.1	Object.....	28
12.2	Characterization	28
12.3	Description	28
12.4	Generation.....	29
13	RQN-M X-ray radiation conditions	29
13.1	Object.....	29
13.2	Characterization	29
13.3	Description	29
13.4	Phantom	30
13.5	Diaphragms	30
13.6	Generation.....	30
14	RQB-M X-ray radiation conditions.....	30
14.1	Object.....	30
14.2	Characterization	31
14.3	Description	31
14.4	Phantom	31
14.5	Diaphragm.....	32
14.6	Generation.....	32
	Annex A (normative) Measuring arrangements.....	33
	Annex B (informative) Determination of the amount of additional filtration.....	39

Annex C (informative) Tabulated values for the squared signal-to-noise ratio per air kerma (SNR_{in2}).....	40
Annex D (normative) Additional X-ray radiation conditions as used in mammography and determination of the corresponding nominal aluminium half-value layers.....	41
D.1 Object.....	41
D.2 Characterization	41
D.3 Description	41
D.4 Generation and verification	41
Annex E (informative) Overview of X-ray radiation conditions	43
Bibliography.....	44
List of comments.....	45
Figure A.1 – Measuring arrangement for achieving RQR 2 to RQR 10 X-ray radiation conditions	33
Figure A.2 – Measuring arrangement for achieving RQA 2 to RQA 10 X-ray radiation conditions	34
Figure A.3 – Measuring arrangement for achieving RQN 2 to RQN 10 X-ray radiation conditions	35
Figure A.4 – Measuring arrangement for achieving RQB 2 to RQB 10 X-ray radiation conditions	36
Figure A.5 – Measuring arrangement for achieving RQN-M X-ray radiation conditions	37
Figure A.6 – Measuring arrangement for achieving RQB-M X-ray radiation conditions	38
Figure B.1 – Determination of additional filtration required for adjusting the total filtration to the prescribed value (see 5.4).....	39
Table 1 – Parameters for RQR 2 to RQR 10 X-ray radiation conditions	18
Table 2 – Parameters for RQA 2 to RQA 10 X-ray radiation conditions	21
Table 3 – Parameters for RQC 3, RQC 5 and RQC 8 X-ray radiation conditions.....	22
Table 4 – Parameters for RQT 8, RQT 9 and RQT 10 X-ray radiation conditions.....	23
Table 5 – Parameters for RQR-M 1 to RQR-M 4 X-ray radiation conditions	28
Table 6 – Parameters for RQA-M 1 to RQA-M 4 X-ray radiation conditions	29
Table 7 – Parameters for RQN-M 1 to RQN-M 4 X-ray radiation conditions	30
Table 8 – Parameters for RQB-M 1 to RQB-M 4 X-ray radiation conditions	31
Table C.1 – SNR_{in2} values for the RQA X-ray radiation conditions.....	40

INTERNATIONAL ELECTROTECHNICAL COMMISSION

Medical diagnostic X-ray equipment - Radiation conditions for use in the determination of characteristics

FOREWORD

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.

2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.

3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.

4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.

5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.

6) All users should ensure that they have the latest edition of this publication.

7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.

8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.

9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

This commented version (CMV) of the official standard IEC 61267:2025 edition 3.0 provides the user with comments from IEC SC 62C experts explaining the reasons for the most relevant changes made to the previous IEC 61267:2005 edition 2.0, or clarifying specific parts of the content.

Experts' comments are identified by a blue-background number. Mouse over a number to display a pop-up note with the comment.

This publication contains the CMV and the official standard. The full list of comments is available at the end of the CMV.

IEC 61267 has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

This third edition cancels and replaces the second edition published 2005. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) removing former Annex C "Measurement of the practical peak voltage";
- b) inserting informative [Annex C](#) "Tabulated values for the squared signal-to-noise ratio per air kerma (SNR_{in}^2)" and normative [Annex D](#) "Additional X-ray radiation conditions as used in mammography and determination of the corresponding nominal aluminium half-value layers";
- c) revision of X-ray radiation conditions;
- d) new method for verification of X-ray radiation conditions;
- e) change of term definitions.

The text of this International Standard is based on the following documents:

Draft	Report on voting
62C/958/FDIS	62C/965/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

INTRODUCTION

To establish characteristics, aspects or properties of [associated equipment](#) or to have available [radiation beams](#) for physical and medical investigations, sets of well-defined [X-ray radiation conditions](#) can offer an important tool in many situations.

From a regulation and standardization point of view, there is a need

- to have available well-defined [X-ray radiation conditions](#) that can be used internationally to specify standards of operation of [X-ray equipment](#),
- to provide a basis for the harmonization of existing national standards,
- to provide uniform sets of [X-ray radiation conditions](#) (a dictionary of [X-ray radiation conditions](#)) to describe and judge the performance of X-ray equipment for the benefit of [manufacturers](#), [users](#), [patients](#) and health protection authorities, and
- to solve communication problems between [manufacturers](#), [users](#) and regulatory authorities, stemming from a lack of internationally accepted definitions and test methods.

From an application point of view, commonly accepted sets of [X-ray radiation conditions](#) would in general find use in

- [quality control](#) tests by [manufacturers](#),
- installation and [acceptance tests](#),
- calibration of test instrumentation,
- type approval tests (where required),
- inspection and tests by regulatory authorities and testing institutes,
- physical and medical studies in physical laboratories and medical facilities, and
- determination of characteristics of [associated equipment](#).

Standardized [X-ray radiation conditions](#) can benefit a range of potential [users](#), such as

- [manufacturers](#) of [X-ray equipment](#),
- [manufacturers](#) of X-ray test instrumentation,
- research laboratories,
- testing institutes,
- government regulatory authorities,
- service organizations, and
- standardization organizations.

The [X-ray radiation conditions](#) defined in this document are intended to represent the range of typical X-ray [radiation beams](#) encountered in medical diagnostic X-ray equipment. This includes X-ray [radiation beams](#) passing through the filtration of an [X-ray source assembly](#) whereby the [radiation field](#) includes only an insignificant amount of [scattered radiation](#). It also includes the more general case, where [scattered radiation](#) emerges from an [exit surface](#) of a [patient](#) or a [phantom](#). An overview of the [X-ray radiation conditions](#) defined in this document and of possible applications can be found in [Annex E](#).

Potential applications include studies for devices used in specific imaging modalities such as mammography. However, the clauses of this document are not intended to represent specific imaging modalities in general. For example, the [X-ray radiation conditions](#) described in [Clause 5](#) can be useful for examinations of equipment found in dental radiography but also for examinations of equipment related to chest radiography. In addition, some [X-ray radiation conditions](#) can only partially cover the range of equipment for a particular imaging task. Therefore, imaging modalities are not explicitly included or excluded from the scope of this document.

1 Scope

This document applies to test procedures which, for the determination of characteristics of systems or components of medical diagnostic **X-ray equipment**, require well-defined **X-ray radiation conditions**.

This document deals with methods for generating **X-ray radiation conditions** which can be used under test conditions typically found in test laboratories or in manufacturing facilities for the determination of characteristics of medical diagnostic **X-ray equipment**.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 61674, *Medical electrical equipment - Dosimeters with ionization chambers and/or semiconductor detectors as used in X-ray diagnostic imaging*

IEC 61676, *Medical electrical equipment - Dosimetric instruments used for non-invasive measurement of X-ray tube voltage in diagnostic radiology*

Bibliography

- [1] IEC 60731:2011, *Medical electrical equipment - Dosimeters with ionization chambers as used in radiotherapy*
- [2] IEC 61676:2023, *Medical electrical equipment - Dosimetric instruments used for non-invasive measurement of X-ray tube voltage in diagnostic radiology*
- [3] IEC TR 60788:2004, *Medical electrical equipment - Glossary of defined terms*
- [4] ICRU Report 60, *Fundamental Quantities and Units for Ionizing Radiation*
- [5] IEC 61267:2005, *Medical diagnostic X-ray equipment - Radiation conditions for use in the determination of characteristics*
- [6] IEC 62220-1-1, *Medical electrical equipment - Characteristics of digital X-ray imaging devices - Part 1-1: Determination of the detective quantum efficiency - Detectors used in radiographic imaging*
- [7] IEC 62220-1-3, *Medical electrical equipment - Characteristics of digital X-ray imaging devices - Part 1-3: Determination of the detective quantum efficiency - Detectors used in dynamic imaging*
- [8] ISO/IEC Guide 98-3:2008/Suppl 1:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995) — Supplement 1: Propagation of distributions using a Monte Carlo method*
- [9] Stefan Pojtinger, On the robustness of detective quantum efficiency within the limits of IEC 61267 RQA standard radiation qualities, *Radiation Protection Dosimetry, Radiat. Prot. Dosim.* Volume 200, Issue 5, Pages 515–522
- [10] Steffen Ketelhut et al, Catalog of x-ray spectra of Mo-, Rh-, and W-anode-based x-ray tubes from 10 to 50 kV, *Phys. Med. Biol.* 66 115013