



IEC 62127-3

Edition 2.1 2026-02

INTERNATIONAL STANDARD

CONSOLIDATED VERSION

**Ultrasonics - Hydrophones -
Part 3: Properties of hydrophones for ultrasonic fields**

CONTENTS

FOREWORD	3
INTRODUCTION.....	5
INTRODUCTION to Amendment 1	5
1 Scope.....	6
2 Normative references	6
3 Terms, definitions and symbols	6
4 List of symbols	9
5 Hydrophone characteristics	10
5.1 General.....	10
5.2 Basic information	10
5.3 Sensitivity	11
5.4 Frequency response	11
5.4.1 Stated frequency band.....	11
5.4.2 Frequency dependence	12
5.5 Directional response	12
5.5.1 General	12
5.5.2 Determination of the directional response.....	12
5.5.3 Symmetry of directional response	13
5.6 Effective hydrophone size	13
5.6.1 General	13
5.6.2 Model of directional response	14
5.6.3 Fitting of experimental response to theoretical predictions.....	15
5.7 Dynamic range, linearity and electromagnetic interference.....	16
5.8 Electric output characteristics	16
5.8.1 Hydrophone without pre-amplifier	16
5.8.2 Hydrophone assembly	17
5.8.3 Output lead configuration.....	17
5.9 Environmental aspects.....	17
5.9.1 Temperature range	17
5.9.2 Water tightness	18
5.9.3 Water properties and incompatible materials	18
5.9.4 Exposed material.....	18
5.10 Guidance manual.....	18
5.11 List of hydrophone characteristics.....	18
Annex A (informative) Examples of information on hydrophone properties	19
A.1 General.....	19
A.2 Basic information	19
A.3 Sensitivity and frequency response.....	19
A.4 Directional response	23
A.5 Effective dimension.....	24
A.6 Dynamic range, linearity and electromagnetic interference.....	25
A.6.1 Lower dynamic limit	25
A.6.2 Upper dynamic limit	25
A.7 Electric output characteristics	26
A.8 Environmental aspects.....	26
Annex B (informative) Rationale.....	27

B.1	General.....	27
B.2	Changes to the determination of directional response	27
B.3	Changes to the determination of effective radius.....	27
Annex C (informative)	Membrane hydrophone directivity model	29
C.1	General.....	29
C.2	Details of model.....	29
Bibliography	30
Figure A.1	– Frequency response of 0,2 mm needle hydrophone in the range 1 MHz to 40 MHz	21
Figure A.2	– Frequency response of 0,2 mm needle hydrophone in the range 100 kHz to 1 MHz	22
Figure A.3	– Directional response of 0,2 mm needle hydrophone	24
Figure A.4	– Effective radius of 0,2 mm needle hydrophone.....	24
Figure A.5	– Comparison of modelled and experimentally derived directional response	25
Table A.1	– Example of basic information for 0,2 mm needle hydrophone assembly	19
Table A.2	– Frequency response of 0,2 mm needle hydrophone in the range 1 MHz to 40 MHz	20
Table A.3	– Frequency response of 0,2 mm needle hydrophone in the range 100 kHz to 1 MHz	21

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**Ultrasonics - Hydrophones -
Part 3: Properties of hydrophones for ultrasonic fields**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 62127-3 edition 2.1 contains the second edition (2022-12) [documents 87/818/FDIS and 87/824/RVD] and its amendment 1 (2026-02) [documents 87/897/CDV and 87/914/RVC].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

IEC 62127-3 has been prepared by IEC technical committee 87: Ultrasonics. It is an International Standard.

This second edition cancels and replaces the first edition published in 2007 and Amendment 1:2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition.

- a) The upper frequency limit of 40 MHz has been removed.
- b) Hydrophone sensitivity definitions have been changed to recognize sensitivities as complex-valued quantities.
- c) Procedures to determine the effective hydrophone size have been changed according to the rationale outlined in Annex B.
- d) Requirements on the frequencies for which the effective hydrophone size shall be provided have been changed to achieve practicality for increased frequency bands.
- e) The new Annex B and Annex C have been added.
- f) Annex A has been updated to reflect the changes of the normative parts.

The text of this International Standard is based on the following documents:

Draft	Report on voting
87/818/FDIS	87/824/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

A list of all parts of IEC 62127 series, published under the general title *Ultrasonics – Hydrophones*, can be found on the IEC website.

NOTE Words in **bold** in the text are defined in Clause 3.

The committee has decided that the contents of this document and its amendment will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

INTRODUCTION

The spatial and temporal distribution of acoustic pressure in an ultrasonic field in a liquid medium is commonly determined using miniature ultrasonic **hydrophones**. The properties of these **hydrophones** have been dealt with in a number of IEC standards in various aspects. The purpose of this document is to bring together all these specifications and to establish a common standard on the properties of ultrasonic **hydrophones**. The main **hydrophone** application in this context is the measurement of ultrasonic fields emitted by medical diagnostic equipment in water. Other medical applications are field measurements for therapy equipment such as that used in lithotripsy, high-intensity focused ultrasound (HIFU) and physiotherapy. **Hydrophones** are also used extensively in non-medical applications for both product development and quality control including:

- mapping of the ultrasound field within ultrasonic cleaning baths;
- characterization of acoustic fields used in transmission measurement systems (e.g. ultrasonic spectrometers, ultrasonic attenuation meters and velocimeters);
- characterization of acoustic fields used in reflection measurement systems (e.g. Doppler flowmeters).

While the term **hydrophone** can be used in a wider sense, it is understood here as referring to miniature piezoelectric **hydrophones**. It is this instrument type that is used today in various areas of ultrasonics and, in particular, to quantitatively characterize the field structure of medical diagnostic instruments. With regard to other pressure sensor types, such as those based on fibre optics, some of the requirements of this document are applicable to these as well but others are not. If in the future these other **hydrophone** types gain more importance in field measurement practice, their properties will have to be dealt with in a revised version of this document or in a separate one.

Underwater **hydrophones** as covered by IEC 60500, IEC 60565-1, and IEC 60565-2 are not included in this document, although there is an overlap in the frequency ranges. Underwater **hydrophones** are used in natural waters, even in the ocean, and this leads to different technical concepts and requirements. In addition, the main direction of acoustic incidence in underwater applications is at various angles and often at right angles to the **hydrophone axis**, whereas in this document it is assumed that the main direction of acoustic incidence is in the direction of the **hydrophone axis**.

Historically, ultrasonic **hydrophones** were used almost exclusively as amplitude sensors. However, the complex-valued nature of a **hydrophone's** system response function is well understood and IEC 62127-1:2022 makes use of this within the deconvolution procedures it contains. In this document, requirements are specified for the amplitude aspect of the **hydrophone** sensitivity and recommendations are provided for the phase aspect which can be derived either via calibration, or via calculation methods that are discussed in IEC 62127-1:2022.

INTRODUCTION to Amendment 1

The second edition of IEC 62127-3 was published in 2022. The application of the document has revealed improvement potential with respect to the measurement requirements for **directional responses** and regarding data evaluation requirements for **effective hydrophone size** determination. This amendment addresses these items to ensure the practicality of the required measurements as well as the generation of a common data base to support **hydrophone** spatial averaging correction methods and the simplifying assumptions therein.

1 Scope

This part of IEC 62127 specifies relevant **hydrophone** characteristics.

This document is applicable to:

- **hydrophones** employing piezoelectric sensor elements, designed to measure the pulsed and continuous wave ultrasonic fields generated by ultrasonic equipment;
- **hydrophones** used for measurements made in water;
- **hydrophones** with or without an associated pre-amplifier.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 62127-1, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields*

IEC 62127-2, *Ultrasonics – Hydrophones – Part 2: Calibration for ultrasonic fields* ~~up to 40 MHz~~

Bibliography

IEC documents

IEC 60050-801, *International Electrotechnical Vocabulary – Part 801: Acoustics and electroacoustics*

IEC 60050-802, *International Electrotechnical Vocabulary – Part 802: Ultrasonics*

IEC 60565-1, *Underwater acoustics – Hydrophones – Calibration of hydrophones – Part 1: Procedures for free-field calibration of hydrophones*

IEC 60565-2, *Underwater acoustics – Hydrophones – Calibration of hydrophones – Part 2: Procedures for low frequency pressure calibration*

Other references

- [1] RADULESCU, EG., LEWIN, PA., NOWICKI, A. and BERGER, WA. Hydrophones' effective diameter measurements as a quasi-continuous function of frequency. *Ultrasonics*, 2003, vol. 41, iss. 8, p. 635-641
- [2] ISO/IEC Guide 98-3:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*
- [3] FAY, B., LEWIN, PA., LUDWIG, G., SESSLER, GM. and YANG, G. The influence of spatial polarization distribution on spot poled PVDF membrane hydrophone performance. *Ultrasound Med. Biol.*, 1992, vol. 18, no. 6-7, p. 625-635
- [4] SHOMBERT, DG. SMITH, SW. and HARRIS, GR. Angular response of miniature ultrasonic hydrophone. *Med. Phys.* 1982, vol 9, no. 4, p. 484-492
- [5] KRÜCKER, JF., EISENBERG, AE., KRIX M., LÖTSCH R., PESSEL M., and TRIER H-G. Rigid piston approximation for computing the transfer function and angular response of a fiber-optic hydrophone. *J. Acoust. Soc. Am.*, 2004, vol. 107, no. 4, p. 1994-2003
- [6] BACON, DR. Characteristics of a pvdf membrane hydrophone for use in the range 1-100 MHz. *IEEE Trans. SU*. 1982, vol 29, no. 1, p. 18-25
- [7] SMITH, RA. and BACON, DR. A multiple-frequency hydrophone calibration technique. *J. Acoust. Soc. Am.*, 1990, vol. 87, p. 2231-2243
- [8] FAY, B., LUDWIG, G., LANKJAER, C. and LEWIN, PA. Frequency response of PVDF needle-type hydrophones. *Ultrasound Med. Biol.* 1994, vol. 20, no. 4, p. 361-366
- [9] AIUM/NEMA. *Acoustic output measurement standard for diagnostic ultrasound equipment*. (NEMA Standards Publication UD 2-2004, Revision 3). Laurel, MD: American Institute of Ultrasound in Medicine (AIUM); Rosslyn, VA: National Electrical Manufacturers Association (NEMA), 2004
- [10] WILKENS V. and MOLKENSTRUCK W. *Broadband PVDF membrane hydrophone for comparisons of calibration methods up to 140 MHz*, IEEE Trans. UFFC. 2007, vol. 54, pp. 1784-1791