

TECHNICAL REPORT

**Safety of laser products –
Part 8: Guidelines for the safe use of lasers on humans**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.99; 31.260

ISBN 978-2-8322-5285-7

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
1 Scope.....	8
2 Normative references	8
3 Terms and definitions	8
4 Hazards and preventive measures.....	10
4.1 Risks to eyes	10
4.1.1 General	10
4.1.2 Laser protective eyewear of personnel.....	10
4.1.3 Laser protective eyewear of patients or clients	11
4.1.4 Eye protection with viewing optics	11
4.1.5 Eye protection of persons behind room windows.....	11
4.1.6 Reflecting surfaces	12
4.2 Risks to skin	12
4.2.1 General	12
4.2.2 Skin protection against laser radiation	13
4.2.3 Protection against burning of materials close to the skin.....	13
4.3 Risk of internal combustion	13
4.3.1 General	13
4.3.2 Protection against combustion of endogenous gases	13
4.3.3 Protection against airway fire.....	14
4.3.4 Protection against burning of an endoscope	14
4.4 Risks due to inhalation of noxious fumes and plumes.....	14
4.4.1 General	14
4.4.2 Dedicated smoke evacuation systems	15
4.4.3 Central vacuum suction systems.....	15
4.4.4 Face masks	15
4.4.5 Disposal of filters.....	15
5 Administrative procedures	16
5.1 LASER SAFETY OFFICER (LSO).....	16
5.1.1 General	16
5.1.2 Duties and responsibilities of the LSO	16
5.2 INCIDENTS and ACCIDENTS	17
5.2.1 General	17
5.2.2 Initial reporting	17
5.2.3 Medical examination	17
5.2.4 Medical surveillance	17
5.2.5 Investigation of the circumstances of the event.....	17
5.2.6 INCIDENT and ACCIDENT follow-up	18
5.3 Maintenance and inspection.....	18
5.3.1 Maintenance	18
5.3.2 Inspection schedule	19
6 Laser safety training recommendations.....	19
7 Laser environment.....	19
7.1 The LASER CONTROLLED AREA	19

7.2	Access controls.....	19
7.2.1	General	19
7.2.2	Warning signs.....	19
7.2.3	Illuminated warning indicators.....	20
7.2.4	Door switches and interlocks	20
7.3	Fire protection policy.....	20
Annex A (informative) Biological effects, hazards, laser equipment technology		21
A.1	Biological effects and hazards	21
A.1.1	General	21
A.1.2	Hazards to the eye	22
A.1.3	Hazards to the skin.....	23
A.1.4	Determination of NOHD	24
A.2	Laser applications.....	24
A.2.1	General	24
A.2.2	Area of application.....	24
A.2.3	Types of application.....	24
A.3	Laser equipment technology	25
A.3.1	Laser radiation sources	25
A.3.2	Laser radiation delivery systems.....	25
Annex B (informative) Window shielding		29
B.1	General.....	29
B.2	Laser wavelength.....	29
B.3	Resistance to fire and heat	29
B.4	Removable attachments.....	29
Annex C (informative) Checklist for laser installation.....		31
C.1	General.....	31
C.2	Identify.....	31
C.3	Determine relevant information	31
C.3.1	Details of laser	31
C.3.2	Other hazards.....	31
C.3.3	Application.....	31
C.3.4	Life cycle	31
C.4	Risk assessment.....	32
C.4.1	General	32
C.4.2	Laser beam	32
C.4.3	Non-laser hazards	32
C.5	Treatment unit	32
C.6	Authorization and training of personnel	33
C.7	Operating procedures	33
C.7.1	Pre-use testing	33
C.7.2	INCIDENT procedure.....	33
C.8	Annual audit.....	33
C.8.1	Installation.....	33
C.8.2	Risk assessment	33
Annex D (informative) Laser safety training.....		34
Annex E (informative) Inspection schedule		35
E.1	General.....	35
E.2	Quality assurance (QA) tests	35

E.2.1	General	35
E.2.2	Cables	35
E.2.3	Emergency switches	36
E.2.4	Interlocks.....	36
E.2.5	Indicators	36
E.2.6	Beam RADIANT POWER or pulse energy	36
E.2.7	Articulated arm	36
E.2.8	Beam coincidence	36
E.2.9	Optical fibres	36
E.2.10	Aiming beam.....	37
E.2.11	Calibration of the RADIANT POWER (see also E.2.6)	37
E.2.12	Specialized accessories	37
E.2.13	Protective eyewear	37
E.3	Preventive maintenance	37
E.3.1	General	37
E.3.2	Cleaning and disinfection.....	38
E.3.3	Preventive maintenance checklist.....	38
E.3.4	Checks prior to use.....	38
Annex F (informative)	Safety issues in laser applications	40
F.1	General.....	40
F.2	Use of optical fibres	40
F.2.1	General	40
F.2.2	Hazard summary	40
F.2.3	Precautions	40
F.3	Use of flexible endoscopes	41
F.3.1	General	41
F.3.2	Hazard summary	41
F.3.3	Precautions	41
F.4	Use of lasers with rigid endoscopes, microscopes and colposcopes.....	42
F.4.1	General	42
F.4.2	Hazard summary	42
F.4.3	Precautions	42
F.5	Use of lasers with free hand manipulation capability	42
F.5.1	General	42
F.5.2	Hazard summary	42
F.5.3	Precautions	43
F.6	Use of lasers in eye surgery.....	43
F.6.1	Hazard summary	43
F.6.2	Precautions	43
F.7	Use of lasers in conjunction with anaesthesia	43
F.7.1	Hazard summary	43
F.7.2	Precautions	43
Bibliography	45
Table A.1	– Summary of pathological effects associated with excessive exposure to light	22
Table D.1	– List of training items and allocation to persons involved (arbitrary order)	34
Table E.1	– Inspection schedule	35

INTERNATIONAL ELECTROTECHNICAL COMMISSION

SAFETY OF LASER PRODUCTS –**Part 8: Guidelines for the safe use of lasers on humans**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC TR 60825-8 has been prepared by IEC technical committee 76: Optical radiation safety and laser equipment. It is a Technical Report.

This third edition cancels and replaces the second edition published in 2006. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition.

- a) Recent medical laser equipment classified as laser class 1C is now included. Equipment of laser class 1C incorporates sensors which ensure good contact, so that laser emission into free space is inhibited.
- b) More emphasis is given to protective eyewear of patients or clients, to the burning of materials close to the skin and to collateral hazards such as from internal or external fire and from noxious gases.
- c) General technical update.

The text of this Technical Report is based on the following documents:

Draft	Report on voting
76/640/DTR	76/658/RVDTR

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this Technical Report is English.

Terms written in small capitals in this document are defined in Clause 3.

A list of all parts in the IEC 60825 series, published under the general title *Safety of laser products*, can be found on the IEC website.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

Lasers emit visible or invisible optical radiation or both. In some cases, this radiation is a parallel beam with almost no divergence. This means that the inherently high IRRADIANCE of the laser is maintained over considerable distances. Due to the laser irradiation properties, injuries to the eye and skin can occur. Annex A includes descriptions of laser systems and some medical applications.

Lasers present hazards to anyone present during the operation of the laser. Serious risks of injury, particularly to the eye, or undesired effects can result from lack of protective measures, the use of faulty laser equipment, misdirected beams or inappropriate laser controls or settings.

Lasers which are used in contact mode on the skin may be classified as laser class 1C. These laser systems incorporate safety means which ensure that laser radiation can only be emitted if the interlocks detect good contact with the skin so that free space emission above the AEL of class 1 is prohibited. When used correctly, class 1C laser systems are considered safe for the eyes.

This document is intended to give direction as to how aspects of laser safety are incorporated into medical laser practice. It is not intended to take precedence over existing or proposed national guidance. However, where none exists, this document is intended to provide helpful information.

Although the LASER USER has direct responsibility for safety during laser use, the employer, referred to in this document as RESPONSIBLE PERSON, bears the responsibility for the setting up of a framework for the safe use of the system. A LASER SAFETY OFFICER (LSO) can be appointed to provide expert advice to the RESPONSIBLE PERSON and all personnel concerned with the laser operation. This document emphasizes the need for appropriate laser safety training for all staff involved in providing practical guidance on installation, operation, maintenance and servicing.

SAFETY OF LASER PRODUCTS –

Part 8: Guidelines for the safe use of lasers on humans

1 Scope

This part of IEC 60825, which is a Technical Report, serves as a guide to the employer, the RESPONSIBLE PERSON, the LASER SAFETY OFFICER, the LASER USER and other persons involved, on the safe use of lasers and laser equipment classified as laser class 1C, 3B or 4 in interventional applications of laser beams on humans, excluding use of consumer products.

NOTE Premises where lasers are used include, but are not limited to, health-care facilities, dental-care practices, physiotherapy, beauty-care and cosmetic facilities.

This document explains the control measures recommended for the safety of the LASER USER, patients, clients, staff, maintenance personnel and others. Engineering controls which form part of the laser equipment or the installation are also briefly described to provide an understanding of the general principles of protection.

The subject areas covered in this document include

- BEAM DELIVERY SYSTEMS;
- biological effects of laser radiation;
- reporting of ACCIDENTS and dangerous situations, and
- checklists.

The object of this document is to enhance the protection of persons from laser radiation and other associated hazards by providing guidance on how to establish safety procedures, precautions and user control measures.

Medically relevant advice such as about treatment indications, counter-indications, patient or client condition, medical or beauty-care treatment procedures, patch testing, medication, adverse tissue or skin conditions and follow-up controls is beyond the scope of this document.

2 Normative references

There are no normative references in this document.