



TECHNICAL REPORT

Nuclear medicine instrumentation – Routine tests – Part 3: Positron emission tomographs

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.50

ISBN 978-2-8322-5230-7

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	3
1 Scope.....	5
2 Normative references	5
3 Terms and definitions	5
4 Test methods.....	8
4.1 RELATIVE SENSITIVITY PER LINE OF RESPONSE and accuracy of NORMALIZATION.....	8
4.2 CALIBRATION FACTOR and cross-calibration.....	8
4.3 TRANSVERSE RESOLUTION	9
4.4 Image quality	9
4.5 PET/CT co-registration	9
5 Frequency of ROUTINE TESTS	9
Bibliography.....	10
Index of defined terms	11
Table 1 – Frequency of ROUTINE TESTS	9

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**NUCLEAR MEDICINE INSTRUMENTATION –
ROUTINE TESTS –****Part 3: Positron emission tomographs**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a technical report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC TR 61948-3, which is a technical report, has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 2005. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) a clause to check routine performance tests has been added,

- b) a test to check the accuracy of co-registration of PET and CT images has been added,
- c) a test to check image quality has been added,
- d) the test to check pixel size has been removed.

The text of this technical report is based on the following documents:

Enquiry draft	Report on voting
62C/694/DTR	62C/708/RVDTR

Full information on the voting for the approval of this technical report can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 61948 series, published under the general title *Nuclear medicine Instrumentation – Routine tests*, can be found on the IEC website.

Terms used throughout this document that have been defined in Clause 3 appear in SMALL CAPITALS.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

NUCLEAR MEDICINE INSTRUMENTATION – ROUTINE TESTS –

Part 3: Positron emission tomographs

1 Scope

This part of IEC 61948 describes test methods for POSITRON EMISSION TOMOGRAPHS (PET). As part of QUALITY CONTROL, this document is defining ROUTINE TESTS to be performed by the user of POSITRON EMISSION TOMOGRAPHS to maintain proper operation conditions. The results of these ROUTINE TESTS are compared to the REFERENCE DATA determined during or after ACCEPTANCE TEST. Methods used for ACCEPTANCE TESTS are described in IEC 61675-1:2013.

In addition, today a POSITRON EMISSION TOMOGRAPH often includes X-RAY EQUIPMENT for COMPUTED TOMOGRAPHY (CT). For this document, PET/CT hybrid devices are considered to be state of the art, dedicated POSITRON EMISSION TOMOGRAPHS not including the X-ray component being special cases only.

QUALITY CONTROL tests specific to only the CT component of the PET/CT are described in IEC 61223-2-6. The CT SCANNER also is subject to a TYPE TEST according to IEC 60601-1 and applicable collateral and particular standards.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 61223-2-6:2006, *Evaluation and routine testing in medical imaging departments – Part 2-6: Constancy tests – Imaging performance of computed tomography X-ray equipment*

IEC 61675-1:2013, *Radionuclide imaging devices – Characteristics and test conditions – Part 1: Positron emission tomographs*

IEC TR 61948-1:2016, *Nuclear medicine instrumentation – Routine tests – Part 1: Gamma radiation counting systems*