



IEC/TR 62649

Edition 1.0 2010-04

TECHNICAL REPORT

**Requirements for measurement standards for high intensity therapeutic
ultrasound (HITU) devices**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

PRICE CODE **XD**

ICS 17.140.50

ISBN 978-2-88910-922-7

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Background.....	7
3 Limitations of existing standard methods.....	10
3.1 General.....	10
3.2 Very high pressures.....	11
3.3 Very high intensities.....	11
3.4 Strong focusing.....	12
3.5 Nonlinear harmonics.....	12
3.6 Acoustic saturation and nonlinear loss.....	12
3.7 Relevant parameters.....	13
4 Survey of experts.....	13
5 Existing literature for measurement of HITU fields.....	18
6 Discussion.....	26
6.1 General.....	26
6.2 Measurement of power.....	26
6.3 Specification of field parameters related pressure and intensity distribution.....	27
6.4 Robust methodology for measuring pressure at a point.....	27
6.5 Direct measurement of intensity.....	28
6.6 Measurement of temperature rise and temperature distribution.....	28
6.7 Thermal dose.....	28
6.8 Cavitation activity.....	30
6.9 Registration of the HITU field with the targeting system.....	30
6.10 Tissue-mimicking material for QA/engineering evaluation.....	30
6.11 Electrical properties of the transducer.....	31
6.12 Treatment monitoring.....	31
6.13 Equipment safety and essential performance.....	31
6.14 Tissue properties.....	31
7 Recommendations.....	32
7.1 Items for immediate development.....	32
7.1.1 General.....	32
7.1.2 Measurement of total output power as an International Standard or as an amendment to IEC 61161 Ed.2 in TC 87.....	32
7.1.3 Specification and measurement of field parameters as a Technical Specification (TS) in TC 87.....	32
7.1.4 Equipment safety and essential performance as a Particular Standard in the 60601 series in TC 62.....	32
7.2 Items for development within 5 years.....	32
7.2.1 General.....	32
7.2.2 Robust method of measuring pressure.....	32
7.2.3 Measurement of temperature.....	33
7.2.4 Electrical properties of therapeutic transducers.....	33
7.2.5 Registration of the HITU field with the targeting system.....	33
7.2.6 Tissue-mimicking material for QA/engineering evaluation.....	33
7.3 Items requiring extensive further research.....	33

Annex A (informative) Detailed responses to questionnaire.....	34
Annex B (informative) Definitions in selected ultrasound standards.....	54
Annex C (informative) Chinese national standard.....	67
Bibliography.....	87
Figure 1 – Radiation force balance system using absorbing target (upward beam).....	77
Figure 2 – Radiation force balance system using different measuring system (downward beam)	78
Figure 3 – Hydrophone scanning configuration for measuring HIFU acoustic field (upward beam).....	78
Figure 4 – Hydrophone scanning measurement configuration for measuring HIFU acoustic field (downward beam).....	79
Figure 5 – Correction factor of plane wave for the acoustic field of a circular plane piston ultrasonic transducer	82
Table 1 – General characteristics of ultrasound used for different therapeutic applications	8
Table 2 – International standards and related documents for the measurement of medical ultrasound fields	9
Table 3 – Questions and summarised answers	14

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**REQUIREMENTS FOR MEASUREMENT STANDARDS
FOR HIGH INTENSITY THERAPEUTIC ULTRASOUND (HITU) DEVICES**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a technical report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 62649, which is a technical report, has been prepared by committee TC 87: Ultrasonics.

The text of this technical report is based on the following documents:

Enquiry draft	Report on voting
87/420/DTR	87/428/RVC

Full information on the voting for the approval of this technical report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

INTRODUCTION

This Technical Report is concerned with standards for high intensity therapeutic ultrasound (HITU) and concentrates on applications that destroy tissue by heating which may or may not be accompanied by acoustic cavitation and other mechanisms. The purpose of the report is to identify topics where there is a consensus that the development of international standards would benefit the industries and/or patients involved with these forms of therapeutic ultrasound. The shortcomings of existing standards as they may be related to the applications of interest are reviewed. It is not its purpose to propose or evaluate specific alternative measurement methods which may be more reliably applied to HITU or other therapeutic equipment. Physiotherapy and lithotripsy are excluded as there are existing standards for these established uses. Lower intensity applications such as enhanced bone healing or ultrasound-induced gene therapy are not explicitly considered.

The use of HITU has advanced to the point where systems have achieved clinical approval for general use in several countries. Medical applications and product development are continuing rapidly. The corresponding products of many companies have been approved for marketing and clinical applications. Fast development in preclinical medicine, clinic medicine, and product manufacture has created an urgent need to standardize measurements of the basic acoustic parameters and the field characteristics of HITU. In order to promote the further development of HITU and to ensure its safe and effective use, international standards are required.

REQUIREMENTS FOR MEASUREMENT STANDARDS FOR HIGH INTENSITY THERAPEUTIC ULTRASOUND (HITU) DEVICES

1 Scope

This technical report is relevant to the measurement and specification of ultrasound fields intended for medical therapeutic purposes. Lithotripsy and physiotherapy are excluded, since there are existing International Standards for these applications.

It establishes:

- topics where there is a consensus that the development of International Standards would benefit the industries and/or patients;
- topics where the writing of standards should start immediately;
- topics where the writing of technical specifications should start immediately in order to gain practical experience and establish consensus prior to standardisation;
- topics which require future standardisation but where further research is required before initiating the writing of standards or technical specifications.

This report addresses primarily the requirements for measurement standards related to high intensity therapeutic ultrasound (HITU) [also known as high intensity focused ultrasound (HIFU)] fields which are both high intensity and focused and where the main mechanism for action is thermal. However, aspects of the discussion, conclusions and any resulting standards or technical specifications may also be relevant to therapeutic applications which are either focused or high intensity or where the main mechanism is not thermal.

Scientific literature has been reviewed and responses to a questionnaire which was sent to experts around the world are reported.