

# **IEC TR 62809**

Edition 2.0 2019-10 REDLINE VERSION

# TECHNICAL REPORT



Summary of requirements and tests to products in the scope of IEC 60601-2-66

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ICS 11.180.15; 17.140.50

ISBN 978-2-8322-7556-6

Warning! Make sure that you obtained this publication from an authorized distributor.

# CONTENTS

FOF	REWORD	3
	RODUCTION	
	Scope	
	Normative references	
3	Terms and definitions	6
4	Summary of requirements and tests	7
Bibliography		19
Tab	le 1 – Summary of requirements and tests	8

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

\_\_\_\_\_

# SUMMARY OF REQUIREMENTS AND TESTS FOR PRODUCTS IN THE SCOPE OF IEC 60601-2-66

## **FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This redline version of the official IEC Standard allows the user to identify the changes made to the previous edition. A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text.

**-4-**

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a Technical Report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 62809, which is a Technical Report, has been prepared by IEC technical committee 29: Electroacoustics.

This second edition cancels and replaces the first edition published in 2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) introduction of the term FITTED OSPL90 (FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL) (201.3.206 of IEC 60601-2-66:2019);
- b) the allowable maximum output sound pressure level is now based on FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL (201.9.6 of IEC 60601-2-66:2019).
- c) ESSENTIAL PERFORMANCE is based on risk analysis (201.4.3).

The text of this Technical Report is based on the following documents:

Draft TR	Report on voting
29/1015/DTR	29/1019/RVDTR

Full information on the voting for the approval of this Technical Report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

## INTRODUCTION

During the preparation of IEC 60601-2-66, members of the involved technical committee and working group voiced concerns about the complexity of the document and its structure as part of the IEC 60601 series. Members felt distracted from the technical content by this complexity during reviews of the document stages. There was also concern that groups in the hearing instrument aid community would have problems to understand and apply the standard and that this could be an issue with its acceptability.

In order to have a broad consensus for the new standard, it was agreed that the standard should be supported by this Technical Report, which should enable members of the community and the industry to have a basic understanding of the requirements of the standard, without the need to study the complete standard document and the documents that are referenced in it.

IEC 60601-2-66 was published to address the specific requirements for safety of hearing instruments aids, and it is entitled "Particular requirements for the basic safety and essential performance of hearing-instruments aids and hearing-instruments aid systems". It was published because IEC 60601-1 is a general standard intended to address a wide range of medical electrical equipment – including large scale facilities such as MRI machines, for example – and thus has large sections that are not relevant to low-voltage, low power, subminiature hearing instruments aids.

If IEC 60601-2-66 was not published, test and regulatory organizations would probably have difficulty applying IEC 60601-1, because it does not contain specific guidance for hearing instruments aids. This Technical Report contains all the requirements from IEC 60601-2-66 which relate to hearing instruments aids and reduces discussion with those that do not relate to hearing instruments aids.

It includes specific references to the applicable requirements within IEC 60601-1, and it is suggested that hearing—instrument aid designers and manufacturers along with test and regulatory organizations read this Technical Report as an overview of IEC 60601-2-66.

# SUMMARY OF REQUIREMENTS AND TESTS FOR PRODUCTS IN THE SCOPE OF IEC 60601-2-66

## 1 Scope

This document, which is a Technical Report, provides an overview of the requirements and tests of IEC 60601-2-66 in combination with the applicable sections of IEC 60601-1, and the collateral standards of the IEC 60601 series.

NOTE The IEC 60601 series consists of three levels of standards: IEC 60601-1, known as the general standard, several IEC 60601-1-X documents, known as the collateral standards, and a series of particular standards covering requirements for specific types of equipment (IEC 60601-2-X).

It is intended to assist various groups involved in the product lifecycles process – like designers and suppliers – to get an overview of the basic requirements without studying all involved standard documents in detail. The table includes not all but just the more common requirements and tests.

It is crucial to understand that the summary in this document cannot serve as an input for a product requirement specification or as a test plan without consulting the standard document IEC 60601-2-66 itself. This document alone cannot be used to establish or assess compliance to the standard IEC 60601-2-66.

The summary in Table 1 below does not preclude the user from reading the referenced standards in their entirety for a thorough knowledge of the basic safety of hearing instruments aids and hearing instrument aid systems.

#### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance

IEC 60601-1:2005/AMD1:2012

IEC 60601-2-66:2012/2019, Medical electrical equipment – Part 2-66: Particular requirements for the basic safety and essential performance of hearing—instruments aids and hearing instrument aid systems





Edition 2.0 2019-10

# TECHNICAL REPORT

Summary of requirements and tests to products in the scope of IEC 60601-2-66



# CONTENTS

FOF	REWORD	3
	RODUCTION	
	Scope	
	Normative references	
	Terms and definitions	
4	Summary of requirements and tests	7
Bibliography		. 18
Tab	le 1 – Summary of requirements and tests	8

# INTERNATIONAL ELECTROTECHNICAL COMMISSION

# SUMMARY OF REQUIREMENTS AND TESTS FOR PRODUCTS IN THE SCOPE OF IEC 60601-2-66

## **FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a Technical Report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 62809, which is a Technical Report, has been prepared by IEC technical committee 29: Electroacoustics.

This second edition cancels and replaces the first edition published in 2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) introduction of the term FITTED OSPL90 (FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL) (201.3.206 of IEC 60601-2-66:2019);
- b) the allowable maximum output sound pressure level is now based on FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL (201.9.6 of IEC 60601-2-66:2019).
- c) ESSENTIAL PERFORMANCE is based on risk analysis (201.4.3).

The text of this Technical Report is based on the following documents:

Draft TR	Report on voting
29/1015/DTR	29/1019/RVDTR

Full information on the voting for the approval of this Technical Report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- · replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

## INTRODUCTION

During the preparation of IEC 60601-2-66, members of the involved technical committee and working group voiced concerns about the complexity of the document and its structure as part of the IEC 60601 series. Members felt distracted from the technical content by this complexity during reviews of the document stages. There was also concern that groups in the hearing aid community would have problems to understand and apply the standard and that this could be an issue with its acceptability.

In order to have a broad consensus for the new standard, it was agreed that the standard should be supported by this Technical Report, which should enable members of the community and the industry to have a basic understanding of the requirements of the standard, without the need to study the complete standard document and the documents that are referenced in it.

IEC 60601-2-66 was published to address the specific requirements for safety of hearing aids, and it is entitled "Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems". It was published because IEC 60601-1 is a general standard intended to address a wide range of medical electrical equipment — including large scale facilities such as MRI machines, for example — and thus has large sections that are not relevant to low-voltage, low power, subminiature hearing aids.

If IEC 60601-2-66 was not published, test and regulatory organizations would probably have difficulty applying IEC 60601-1, because it does not contain specific guidance for hearing aids. This Technical Report contains all the requirements from IEC 60601-2-66 which relate to hearing aids and reduces discussion with those that do not relate to hearing aids.

It includes specific references to the applicable requirements within IEC 60601-1, and it is suggested that hearing aid designers and manufacturers along with test and regulatory organizations read this Technical Report as an overview of IEC 60601-2-66.

# SUMMARY OF REQUIREMENTS AND TESTS FOR PRODUCTS IN THE SCOPE OF IEC 60601-2-66

## 1 Scope

This document, which is a Technical Report, provides an overview of the requirements and tests of IEC 60601-2-66 in combination with the applicable sections of IEC 60601-1, and the collateral standards of the IEC 60601 series.

NOTE The IEC 60601 series consists of three levels of standards: IEC 60601-1, known as the general standard, several IEC 60601-1-X documents, known as the collateral standards, and a series of particular standards covering requirements for specific types of equipment (IEC 60601-2-X).

It is intended to assist various groups involved in the product lifecycles process – like designers and suppliers – to get an overview of the basic requirements without studying all involved standard documents in detail. The table includes not all but just the more common requirements and tests.

It is crucial to understand that the summary in this document cannot serve as an input for a product requirement specification or as a test plan without consulting IEC 60601-2-66 itself. This document alone cannot be used to establish or assess compliance to IEC 60601-2-66.

The summary in Table 1 below does not preclude the user from reading the referenced standards in their entirety for a thorough knowledge of the basic safety of hearing aids and hearing aid systems.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance IEC 60601-1:2005/AMD1:2012

IEC 60601-2-66:2019, Medical electrical equipment – Part 2-66: Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems