



TECHNICAL SPECIFICATION



**Ultrasonics – Output test – Guidance for the maintenance of ultrasound
physiotherapy systems**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.60;17.140.50

ISBN 978-2-8322-4561-3

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	7
3 Terms and definitions	7
4 Testing regimes.....	9
4.1 Acceptance testing.....	9
4.2 Weekly testing	9
4.3 Annual testing.....	9
5 Performance testing	9
5.1 Acceptance testing.....	9
5.1.1 General	9
5.1.2 Visual inspection	9
5.1.3 Manufacturer’s statement	9
5.1.4 Ultrasonic output test.....	9
5.1.5 Beam uniformity and output test	10
5.1.6 Recording of results of acceptance test	11
5.1.7 Requirements and recommendation.....	11
5.2 Weekly testing	11
5.2.1 General	11
5.2.2 Visual inspection	11
5.2.3 Relative ultrasonic output test.....	11
5.2.4 Beam uniformity and output test	11
5.2.5 Recording of results of weekly testing.....	11
5.2.6 Requirements and recommendation.....	11
5.3 Annual testing.....	11
5.3.1 General	11
5.3.2 Output power test	12
5.3.3 Effective radiating area.....	12
5.3.4 Beam uniformity test.....	12
5.3.5 Pulse duty factor accuracy test	12
5.3.6 Timer accuracy test	13
5.3.7 Recording of results of annual testing.....	13
5.4 Service requirement.....	13
Annex A (informative) Rationale for testing	15
A.1 Acceptance testing.....	15
A.2 Weekly testing	15
A.3 Annual testing.....	15
Annex B (informative) Guidance for testers.....	16
B.1 Ultrasound power meter specifications	16
B.2 Room and water temperature	17
B.3 Water.....	17
B.4 Environmental considerations	17
B.5 Ultrasound power meter checks	17
B.6 Ultrasound power meter testing technique	17

Annex C (informative) Quantitative relative ultrasonic output test using temperature rise	19
Annex D (informative) Quantitative relative ultrasonic output test using calorimetry	21
Annex E (informative) Qualitative test to assess changes of effective radiating area	23
Annex F (informative) Example of weekly test report	25
Annex G (informative) Example of annual test report	26
Bibliography	30
Figure 1 – Several examples of how to prepare a set-up to check the distortion on the water surface due to ultrasound	14
Figure 2 – Set-up where the slight angle of the treatment head to the vertical may improve the image	14
Figure C.1 – Example of a measurement set-up to measure the temperature rise due to ultrasound in absorbing material	20
Figure D.1 – Schematic of equipment used within the calorimeter method for monitoring power output of therapy treatment heads	21
Figure E.1 – Assessing changes of the effective radiating area using a thermochromic phantom	24
Figure E.2 – Examples of different patterns observed during an experiment using a thermochromic phantom made of silicon	24
Figure G.1 – Example of a power calibration graph for two large applicator heads	28
Figure G.2 – Example of a power calibration graph for two small applicator heads	29

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ULTRASONICS – OUTPUT TEST – GUIDANCE FOR THE MAINTENANCE OF ULTRASOUND PHYSIOTHERAPY SYSTEMS

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. In exceptional circumstances, a technical committee may propose the publication of a technical specification when

- the required support cannot be obtained for the publication of an International Standard, despite repeated efforts, or
- the subject is still under technical development or where, for any other reason, there is the future but no immediate possibility of an agreement on an International Standard.

Technical specifications are subject to review within three years of publication to decide whether they can be transformed into International Standards.

IEC TS 62462, which is a Technical Specification, has been prepared by IEC technical committee 87: Ultrasonics.

This second edition cancels and replaces the first edition published in 2007. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- addition of a novel method for periodic testing regarding possible changes of the **effective radiating area** using thermochromic absorbers in a new Annex E;

The text of this Technical Specification is based on the following documents:

Enquiry draft	Report on voting
87/640/DTS	87/647A/RVDTS

Full information on the voting for the approval of this technical specification can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements: in roman type;
- notes: in small roman type;
- words in **bold** in the text are defined in Clause 3.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

The purpose of this document is to establish standard methods for a qualitative check of the performance of **ultrasound** physiotherapy devices during their lifetime, and to provide guidance on calibration requirements and techniques.

To ensure that the **ultrasound** physiotherapy equipment is in an appropriate condition for use, a regular quality check can be performed. This document defines acceptance, weekly and annual checks. The acceptance test checks the delivery of the device and its performance at the start of its lifetime. The weekly check is a simple qualitative check of device operation. In the annual check, in addition to a qualitative check, a quantitative check is defined. Examples are provided of weekly and annual test reports.

This document also gives guidance to the **testers** concerning the measurement of acoustic output.

Annual testing may be performed by a skilled **tester**, e.g. biomedical engineer, medical physicist, medical device service agent, commercial **tester**, test house, national measurement institute or manufacturer.

ULTRASONICS – OUTPUT TEST – GUIDANCE FOR THE MAINTENANCE OF ULTRASOUND PHYSIOTHERAPY SYSTEMS

1 Scope

This document, which is a Technical Specification, describes methods meant to assist users of **ultrasound** physiotherapy systems in checking the performance of such systems. It is applicable primarily to physiotherapists, general medical practitioners, chiropractors, osteopaths, beauty therapists, sports professionals, biomedical engineers, medical physicists, medical device service agents, commercial **testers**, test houses or manufacturers. Typical **ultrasound** physiotherapy systems operate in the range from 0,5 MHz to 5 MHz. Long-wave **ultrasound** therapy machines operating in the frequency range 30 kHz to 0,5 MHz are not covered by this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE The titles of all publications referred to informatively in this document are listed in the Bibliography.

IEC 60601-2-5:2009, *Medical electrical equipment – Part 2-5: Particular requirements for the basic safety and essential performance of ultrasonic physiotherapy equipment*

IEC 61161:2013, *Ultrasonics – Power measurement – Radiation force balances and performance requirements*

IEC 61689:2013, *Ultrasonics – Physiotherapy systems – Field specifications and methods of measurement in the frequency range 0,5 MHz to 5 MHz*