



International
Standard

ISO 80601-2-70

**Medical electrical equipment —
Part 2-70:
Particular requirements for basic
safety and essential performance
of sleep apnoea breathing therapy
equipment**

Appareils électromédicaux —

*Partie 2-70: Exigences particulières pour la sécurité de base
et les performances essentielles de l'équipement de thérapie
respiratoire pour l'apnée du sommeil*

**Third edition
2025-11**



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	vi
Introduction	vii
201.1 Scope, object and related standards	1
201.1.1 Scope	1
201.1.2 Object	2
201.1.3 Collateral standards.....	2
201.1.4 Particular standards.....	3
201.2 Normative references	4
201.3 Terms and definitions	5
201.4 General requirements	20
201.4.3 <i>Essential performance</i>	20
201.4.3.101 Additional requirements for <i>essential performance</i>	20
201.4.6 <i>ME equipment</i> or <i>ME system</i> parts that contact the patient	21
201.5 General requirements for testing of <i>ME equipment</i>	21
201.5.101 Additional requirements for general requirements for testing of <i>ME equipment</i>	21
201.5.101.1 Gas flowrate specifications.....	21
201.5.101.2 <i>Sleep apnoea breathing therapy equipment</i> testing errors	21
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i>	22
201.7 <i>ME equipment</i> identification, marking and documents	22
201.7.1.2 Legibility of <i>markings</i>	22
201.7.2.4.101 Additional requirements for <i>accessories</i>	22
201.7.2.101 Additional requirements for <i>marking</i> on the outside of <i>ME equipment</i> or <i>ME equipment</i> parts	23
201.7.4.3 Units of measurement	23
201.7.9.2 <i>Instructions for use</i>	23
201.7.9.2.1.101 Additional general requirements	24
201.7.9.2.2.101 Additional requirements for warnings and safety notices	24
201.7.9.2.5.101 Additional requirements for <i>ME equipment</i> description.....	25
201.7.9.2.9.101 Additional requirements for operating instructions.....	25
201.7.9.2.12 <i>Cleaning, disinfection, and sterilization</i>	25
201.7.9.2.13.101 Additional requirements for maintenance	25
201.7.9.2.14.101 Additional requirements for <i>accessories</i> , supplementary equipment, used material.....	26
201.7.9.3.1.101 Additional general requirements	26
201.8 Protection against electrical hazards from <i>ME equipment</i>	27
201.9 Protection against mechanical hazards of <i>ME equipment</i> and <i>ME systems</i>	27
201.9.6.2.1.101 Additional requirements for audible acoustic energy	28
201.10 Protection against unwanted and excessive radiation hazards	30
201.11 Protection against excessive temperatures and other hazards	30
201.11.1.2.2 <i>Applied parts</i> not intended to supply heat to a <i>patient</i>	30
201.11.6.6 <i>Cleaning and disinfection</i> of <i>ME equipment</i> or <i>ME system</i>	30

201.11.7	<i>Biocompatibility of ME equipment and ME systems</i>	31
201.11.8	Additional requirements for interruption of the power supply/supply mains to ME equipment.....	32
201.12	Accuracy of controls and instruments and protection against hazardous outputs	32
201.12.1	Accuracy of controls and instruments	32
201.12.1.101	Stability of static <i>airway pressure accuracy</i> (long-term accuracy).....	32
201.12.1.102	Stability of dynamic <i>airway pressure accuracy</i> (short-term accuracy)	34
201.12.1.102.1	CPAP mode.....	34
201.12.1.102.2	Bi-level positive airway pressure mode, pressure stability.....	35
201.12.1.103	Maximum flowrate.....	38
201.12.4	Protection against hazardous output.....	41
201.12.4.101	Measurement of <i>airway pressure</i>	41
201.12.4.102	Maximum limited pressure protection device	41
201.12.4.103	CO ₂ rebreathing.....	42
201.13	Hazardous situations and fault conditions	42
201.14	Programmable electrical medical systems (PEMS)	42
201.15	Construction of ME equipment	43
201.15.101	Mode of operation	43
201.16	ME systems	43
201.17	Electromagnetic compatibility of ME equipment and ME systems	43
201.101	Breathing gas pathway connectors	44
201.101.1	General.....	44
201.101.2	Other named ports.....	44
201.101.2.1	Patient-connection port.....	44
201.101.2.2	Gas output port.....	44
201.101.2.3	Flow-direction-sensitive components.....	45
201.101.2.4	Accessory access connection port	45
201.101.2.5	Oxygen inlet connector	45
201.101.2.5.1	Low pressure.....	45
201.101.2.5.2	High pressure.....	46
201.102	Requirements for the breathing gas pathway and accessories	46
201.102.1	General.....	46
201.102.2	Labelling.....	46
201.102.3	Humidification.....	47
201.102.4	Breathing system filter (BSF).....	47
201.103	Functional connection	47
201.103.1	General.....	47
201.103.2	Functional connection to support remote supervision	47
202	Electromagnetic disturbances — Requirements and tests	48
202.4.3.1	Configurations	48
202.5.2.2.1	Requirements applicable to all ME equipment and ME systems	48
202.8.1.101	Additional general requirements	48
206	Usability	48

211	Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	49
211.4.2.3.1	Continuous operating conditions.....	49
Annex C (informative)	Guide to <i>marking</i> and labelling requirements for <i>ME equipment</i> and <i>ME systems</i>	50
201.C.1	<i>Marking</i> on the outside of <i>ME equipment</i> , <i>ME systems</i> or their parts.....	50
201.C.2	<i>Accompanying documents</i> , general	50
201.C.3	<i>Accompanying documents</i> , <i>instructions for use</i>	51
201.C.4	<i>Accompanying documents</i> , <i>technical description</i>	53
Annex AA (informative)	Particular guidance and rationale.....	55
AA.1	General guidance	55
AA.2	Rationale for particular clauses and subclauses.....	56
AA.2.1 —	201.1.1 — <i>Scope</i>	56
AA.2.2 —	201.3.207 — <i>Auto CPAP</i>	57
AA.2.3 —	201.3.208 — <i>Automatic start/stop function</i>	58
AA.2.4 —	201.3.254 — <i>Ramp mode</i>	58
AA.2.5 —	201.4.3.101 — Additional requirements for <i>essential performance</i>	58
AA.2.6 —	201.4.6 — <i>ME equipment</i> or <i>ME system</i> parts that contact the <i>patient</i>	59
AA.2.7 —	201.5.101.2 — <i>Sleep apnoea breathing therapy equipment</i> testing errors.....	59
AA.2.8 —	201.7.1.2 — <i>Legibility of markings</i>	59
AA.2.9 —	201.7.9.3.1.101 — Additional general requirements	59
AA.2.10 —	201.9.6.2.1.101 — Additional requirements for audible acoustic energy	60
AA.2.11 —	201.11.6 — <i>Cleaning</i> and <i>disinfection</i> of <i>ME equipment</i> or <i>ME system</i>	60
AA.2.12 —	201.12.1 — Accuracy of controls and instruments	63
AA.2.13 —	201.12.1.102.2 — <i>Bi-level positive airway pressure</i> mode, pressure stability.....	63
AA.2.14 —	201.12.1.103 — Maximum flowrate.....	63
— c) 2)	63	
AA.2.15 —	201.12.4.102 — <i>Maximum limited pressure protection device</i>	63
Annex BB (informative)	Data interface requirements.....	65
BB.1	Background and purpose.....	65
BB.2	Data definition	66
Annex CC (informative)	Reference to the IMDRF <i>essential principles</i> and labelling guidances.....	69
Annex DD (informative)	Terminology — alphabetized index of defined terms	73
Bibliography		77

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Medical equipment, software, and systems*, Subcommittee SC D, *Particular medical equipment, software, and systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 80601-2-70:2020), which has been technically revised.

The main changes compared to the previous edition are as follows:

- updated references;
- updated the audible acoustic energy requirements; and
- updated the maximum flowrate test method.

A list of all parts in the ISO 80601 series and the IEC 80601 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Sleep apnoea is a chronic medical condition where the *patient* repeatedly stops breathing during sleep. These episodes typically last 10 s or more and cause the oxygen levels in the blood to drop. It can be caused by obstruction of the upper *airway* (obstructive sleep apnoea or OSA) or by a failure of the brain to initiate a breath (central sleep apnoea).

NOTE *Sleep apnoea breathing therapy equipment* is intended for the treatment of obstructive sleep apnoea and not central sleep apnoea.

Sleep apnoea, if untreated, can cause and worsen other medical conditions, including hypertension, heart failure and diabetes^[25].

Hypopnoea refers to a transient reduction of airflow, often while the *patient* is asleep, that lasts for at least 10 s, shallow breathing. It also results in arousal or can cause oxygen saturation to drop. Hypopnoea is less severe than apnoea. It is commonly due to partial obstruction of the upper *airway*^[23].

Awareness of the *risks* associated with obstructive sleep apnoea has grown significantly. As a result, the use of *sleep apnoea breathing therapy equipment* to treat obstructive sleep apnoea has become common.

This document covers *basic safety* and *essential performance* requirements needed to protect *patients* in the use of this *ME equipment*.

This document covers *sleep apnoea breathing therapy equipment* for *patient* use. ISO 17510 applies to *masks* and *accessories* used to connect *sleep apnoea breathing therapy equipment* to the *patient*. Figure AA.1 shows this diagrammatically.

In this document, the following print types are used:

- Requirements and definitions: roman type
- *Terms defined in clause 3 of the general standard, in this document or as noted: italic type;*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;

In referring to the structure of this document, the term.

- “clause” means one of the four numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.1, 201.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 201.101, 201.102 and 201.102.1 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

For the purposes of this document, the auxiliary verb:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” is used to describe a possibility or capability; and
- “must” is used to express an external constraint.

Annex C contains a guide to the *marking* and labelling requirements in this document.

Medical electrical equipment—

Part 2-70:

Particular requirements for the basic safety and essential performance of sleep apnoea breathing therapy equipment

201.1 Scope, object and related standards

IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 1 applies, except as follows:

NOTE The general standard is IEC 60601-1:2005+AMD1:2012+AMD2:2020.

201.1.1 Scope

NOTE 1 There is guidance or rationale for this subclause contained in Clause AA.2.1.

IEC 60601-1:2005+AMD1:2012+AMD2:2020, 1.1 is replaced by:

This document is applicable to the *basic safety* and *essential performance* of *sleep apnoea breathing therapy equipment*, hereafter referred to as *ME equipment*, intended to alleviate the symptoms of *patients* who suffer from obstructive sleep apnoea by delivering a therapeutic breathing pressure to the respiratory tract of the *patient*. *Sleep apnoea breathing therapy equipment* is intended for use in the *home healthcare environment* by *lay operators* as well as in professional healthcare institutions.

Sleep apnoea breathing therapy equipment is not considered to utilize a *physiologic closed-loop-control system* unless it uses a physiological *patient* variable to adjust the therapy settings.

This document excludes *sleep apnoea breathing therapy equipment* intended for use with neonates.

This document is applicable to *ME equipment* or an *ME system* intended for those *patients* who are not dependent on *artificial ventilation*. This document is not applicable to *ME equipment* or an *ME system* intended for those *patients* who are dependent on *artificial ventilation* such as *patients* with central sleep apnoea.

This document is also applicable to those *accessories* intended by their *manufacturer* to be connected to *sleep apnoea breathing therapy equipment*, where the characteristics of those *accessories* can affect the *basic safety* or *essential performance* of the *sleep apnoea breathing therapy equipment*.

Masks and application *accessories* intended for use during sleep apnoea breathing therapy are additionally addressed by ISO 17510. Refer to Figure AA.1 for items covered further under this document.

If a clause or subclause is specifically intended to be applicable to *ME equipment* only, or to *ME systems* only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to *ME equipment* and to *ME systems*, as relevant.

Hazards inherent in the intended physiological function of *ME equipment* or *ME systems* within the scope of this document are not covered by specific requirements in this document except in 7.2.13 and 8.4.1 of the general standard.

NOTE 2 See also 4.2 of the general standard.

This document does not specify the requirements for:

- ventilators or *accessories* intended for critical care ventilators for ventilator-dependent *patients*, which are given in ISO 80601-2-12.
- ventilators or *accessories* intended for anaesthetic applications, which are given in ISO 80601-2-13.
- ventilators or *accessories* intended for home care ventilators for ventilator-dependent *patients*, which are given in ISO 80601-2-72.
- ventilators or *accessories* intended for emergency and transport, which are given in ISO 80601-2-84.
- ventilators or *accessories* intended for home-care ventilatory support, which are given in ISO 80601-2-79 and ISO 80601-2-80.
- high-frequency ventilators^[23], which are given in ISO 80601-2-87.
- respiratory high flow equipment, which are given in ISO 80601-2-90;

NOTE 3 ISO 80601-2-80 ventilatory support equipment can incorporate high-flow therapy operational mode, but such a mode is only for spontaneously breathing *patients*.

- user-powered resuscitators, which are given in ISO 10651-4;
- gas-powered emergency resuscitators, which are given in ISO 10651-5;
- oxygen therapy constant flow *ME equipment*; and
- cuirass or “iron-lung” *ventilation* equipment.

201.1.2 Object

IEC 60601-1:2005+AMD1:2012+AMD2:2020, 1.2 is replaced by:

The object of this document is to establish particular *basic safety* and *essential performance* requirements for *sleep apnoea breathing therapy equipment* (as defined in 201.3.259).

NOTE 1 This document has been prepared to address the relevant *essential principles*^[20] and labelling principles^[21] guidances of the International Medical Devices Regulators Forum (IMDRF) as indicated in Annex CC.

NOTE 2 This document has been prepared to address the relevant general safety and performance requirements of European regulation (EU) 2017/745^[19].

201.1.3 Collateral standards

IEC 60601-1:2005+AMD1:2012+AMD2:2020, 1.3 applies with the following addition:

IEC 60601-1-2:2014+AMD1:2020 and IEC 60601-1-6:2010+AMD1:2013+AMD2:2020 apply as modified in Clauses 202, 206 and 211 respectively. IEC 60601-1-3 and IEC 60601-1-9 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards specify *basic safety* and *essential performance* requirements, and may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular *ME equipment* under consideration.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1+AMD1:2012+AMD2:2020 is referred to in this document as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this document corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x”, where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this document.

"Addition" means that the text of this document is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 211 for IEC 60601-1-11, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this document taken together.

Where there is no corresponding clause or subclause in this document, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 2 applies, except as follows:

Addition:

ISO 3744:2010, *Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Engineering methods for an essentially free field over a reflecting plane*

ISO 4871:1996, *Acoustics — Declaration and verification of noise emission values of machinery and equipment*

ISO 5356-1:2015, *Anaesthetic and respiratory equipment — Conical connectors — Cones and sockets*

ISO 14937:2009, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ISO 17510:—¹, *Medical devices — Sleep apnoea breathing therapy — Masks and application accessories*

ISO 17664-1:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 1: Critical and semi-critical medical devices*

ISO 17664-2:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 2: Non-critical medical devices*

ISO 18562-1:2024, *Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 1: Evaluation and testing within a risk management process*

ISO 20417:—², *Medical devices — Information to be supplied by the manufacturer*

ISO 23328-1:2003, *Breathing system filters for anaesthetic and respiratory use — Part 1: Salt test method to assess filtration performance*

ISO 23328-2:2002, *Breathing system filters for anaesthetic and respiratory use — Part 2: Non-filtration aspects*

ISO 80369-1:2025, *Small-bore connectors for liquids and gases in healthcare applications — Part 1: General requirements*

ISO 80369-7:2021, *Small-bore connectors for liquids and gases in healthcare applications Part 7: Connectors for intravascular or hypodermic applications*

ISO 80601-2-74:—³, *Medical electrical equipment — Part 2-74: Particular requirements for basic safety and essential performance of respiratory humidifying equipment*

¹ Under preparation. Stage at the time of publication: ISO/FDIS 17510:2025.

² Under preparation. Stage at the time of publication: ISO/FDIS 20417:2025.

IEC Guide 115:2023, *Application of uncertainty of measurement to conformity assessment activities in the electrotechnical sector*

³ Under preparation. Stage at the time of publication: ISO/FDIS 80601-2-74:2025.